

Data Path : Z:\HPCHEM1\BNA E\DATA\BE012718\
 Data File : BE095192.D
 Acq On : 28 Jan 2018 6:16
 Operator : SJ/JU
 Sample : J1221-08
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6A66

Manual Integrations
 APPROVED

Sohil
 1/29/2018 6:22:24 PM

Quant Time: Jan 29 01:02:13 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 29 00:15:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	3465	0.40	ng/ul	0.00
2) Naphthalene-d8	11.27	136	10926	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	7058	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.73	188	14639m	0.40	ng/ul	0.02
16) Chrysene-d12	21.80	240	15197	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	13279m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.82	152	6924	0.39	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	19282	0.34	ng/ul	0.00
Target Compounds				Ovalue		
11) Pentachlorophenol	17.36	266	6277	2.214	ng/ul	95
12) Phenanthrene	17.75	178	2607	0.053	ng/ul#	88
13) Anthracene	17.85	178	1164	0.026	ng/ul#	78
15) Fluoranthene	19.73	202	4968	0.078	ng/ul	84
17) Pyrene	20.08	202	4453m	0.087	ng/ul	
18) Benzo(a)anthracene	21.79	228	1769m	0.037	ng/ul	
19) Chrysene	21.84	228	1714m	0.035	ng/ul	
21) Benzo(b)fluoranthene	23.62	252	2590m	0.055	ng/ul	
22) Benzo(k)fluoranthene	23.65	252	942m	0.022	ng/ul	
26) Benzo(g,h,i)perylene	28.12	276	886m	0.023	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE012718\
Data File : BE095192.D
Acq On : 28 Jan 2018 6:16
Operator : SJ/JU
Sample : J1221-08
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6A66

Manual Integrations
APPROVED

Sohil
1/29/2018 6:22:24 PM

Quant Time: Jan 29 01:02:13 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 29 00:15:32 2018
Response via : Initial Calibration

