

Data Path : Z:\HPCHEM1\BNA E\DATA\BE012718\
 Data File : BE095164.D
 Acq On : 27 Jan 2018 13:36
 Operator : SJ/JU
 Sample : J1221-10
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6D42

Manual Integrations
 APPROVED

Sohil
 1/29/2018 6:19:54 PM

Quant Time: Jan 28 01:43:59 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 03:18:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	2736	0.40	ng/ul	0.00
2) Naphthalene-d8	11.27	136	8507	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	5935	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.73	188	12245m	0.40	ng/ul	0.02
16) Chrysene-d12	21.80	240	12926	0.40	ng/ul	0.00
20) Perylene-d12	24.45	264	11689m	0.40	ng/ul	0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.82	152	5395	0.39	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	16685	0.36	ng/ul	0.00
Target Compounds						
8) Acenaphthene	15.08	153	2916	0.132	ng/ul#	Ovalue 23

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE012718\
Data File : BE095164.D
Acq On : 27 Jan 2018 13:36
Operator : SJ/JU
Sample : J1221-10
Misc :
ALS Vial : 4 Sample Multiplier: 1




Instrument :
BNA_E
ClientSampleId :
F6D42

Manual Integrations
APPROVED

Sohil
1/29/2018 6:19:54 PM

Quant Time: Jan 28 01:43:59 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 27 03:18:43 2018
Response via : Initial Calibration





#8

Acenaphthene

Concen: 0.132 ng/ul

RT: 15.08 min Scan# 1825

Delta R.T. -0.00 min

Lab File: BE095164.D

Acq: 27 Jan 2018 13:36

Instrument :

BNA_E

ClientSampleId :

F6D42

Tot Ion:	153	Resp:	2916
Ion Ratio	Lower	Upper	
153	100		
152	3.7	36.0	54.0#
154	10.5	72.0	108.0#

Manual Integrations
APPROVED

Sohil
1/29/2018 6:19:54 PM

