

Data Path : Z:\HPCHEM1\BNA\_E\DATA\BE012815\  
 Data File : BE089398.D  
 Acq On : 28 Jan 2015 21:10  
 Operator : TP/IZ  
 Sample : SSTDICC005  
 Misc :  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDICC005

Quant Time: Jan 29 01:22:07 2015  
 Quant Method : Z:\HPCHEM1\BNA\_E\METHODS\SIM-BE012815.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jan 29 01:08:21 2015  
 Response via : Initial Calibration


| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 9.58  | 152  | 88960    | 5.00  | ng    | 0.00     |
| 7) Naphthalene-d8           | 12.50 | 136  | 286326   | 5.00  | ng    | 0.00     |
| 13) Acenaphthene-d10        | 12.21 | 164  | 41601    | 5.00  | ng    | 0.00     |
| 17) Phenanthrene-d10        | 19.98 | 188  | 167125   | 5.00  | ng    | 0.00     |
| 20) Chrysene-d12            | 23.84 | 240  | 290825   | 5.00  | ng    | 0.00     |
| 27) Perylene-d12            | 25.55 | 264  | 368965   | 5.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |       |       |          |
| 3) 2-Fluorophenol           | 6.73  | 112  | 134172   | 4.48  | ng    | 0.00     |
| 4) Phenol-d6                | 8.84  | 99   | 156528   | 4.23  | ng    | 0.00     |
| 8) Nitrobenzene-d5          | 10.84 | 82   | 100534   | 4.53  | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol    | 18.58 | 330  | 10960    | 3.74  | ng    | 0.00     |
| 15) 2-Fluorobiphenyl        | 15.13 | 172  | 149366   | 3.73  | ng    | 0.00     |
| 22) Terphenyl-d14           | 22.48 | 244  | 138545   | 5.22  | ng    | 0.00     |
| Target Compounds            |       |      |          |       |       | Qvalue   |
| 2) n-Nitrosodimethylamine   | 3.55  | 42   | 76085    | 5.15  | ng    | # 91     |
| 5) Phenol                   | 8.86  | 94   | 168335   | 4.51  | ng    | 87       |
| 6) bis(2-Chloroethyl)ether  | 9.06  | 93   | 124674   | 4.46  | ng    | # 98     |
| 9) Nitrobenzene             | 10.89 | 77   | 114639   | 4.62  | ng    | 99       |
| 10) 2,4-Dimethylphenol      | 11.82 | 122  | 84516    | 4.11  | ng    | 100      |
| 11) 2,4-Dichlorophenol      | 12.20 | 162  | 66152    | 4.59  | ng    | 86       |
| 12) Hexachlorobutadiene     | 12.89 | 225  | 45913    | 4.82  | ng    | 99       |
| 16) 2,4-Dinitrophenol       | 16.90 | 184  | 4779     | 5.74  | ng    | 87       |
| 18) Hexachlorobenzene       | 19.23 | 284  | 37152    | 4.46  | ng    | 100      |
| 19) Pentachlorophenol       | 19.65 | 266  | 11054    | 7.70  | ng    | 99       |
| 21) Benzidine               | 22.14 | 184  | 139101   | 6.10  | ng    | 98       |
| 23) Benzo(a)anthracene      | 23.83 | 228  | 1347792  | 6.66  | ng    | 98       |
| 24) 3,3'-Dichlorobenzidine  | 23.81 | 252  | 137172   | 7.58  | ng    | 97       |
| 25) Chrysene                | 23.87 | 228  | 1581239  | 4.78  | ng    | 99       |
| 26) Indeno(1,2,3-cd)pyrene  | 27.07 | 276  | 555936   | 4.76  | ng    | # 100    |
| 28) Benzo(b)fluoranthene    | 25.10 | 252  | 318479   | 12.74 | ng    | 99       |
| 29) Benzo(k)fluoranthene    | 25.13 | 252  | 580746   | 10.23 | ng    | 100      |
| 30) Benzo(a)pyrene          | 25.48 | 252  | 384889   | 11.55 | ng    | 99       |
| 31) Dibenzo(a,h)anthracene  | 27.09 | 278  | 421559   | 15.76 | ng    | 96       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_E\DATA\BE012815\  
Data File : BE089398.D  
Acq On : 28 Jan 2015 21:10  
Operator : TP/IZ  
Sample : SSTDICC005  
Misc :  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDICC005

Quant Time: Jan 29 01:22:07 2015  
Quant Method : Z:\HPCHEM1\BNA\_E\METHODS\SIM-BE012815.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Jan 29 01:08:21 2015  
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 9.58 min Scan# 1132

Delta R.T. 0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10

Instrument :

BNA\_E

LabSampleId :

SSTDICC005


Tgt Ion: 152 Resp: 88960

Ion Ratio Lower Upper

152 100

150 155.8 127.2 190.8




115 58.7 49.3 73.9



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

n-Nitrosodimethylamine

Concen: 5.15 ng

RT: 3.55 min Scan# 317

Delta R.T. -0.03 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10


Tgt Ion: 42 Resp: 76085

Ion Ratio Lower Upper

42 100

74 128.6 101.2 151.8



44 12.0 26.4 39.6#




Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0









#3  
 2-Fluorophenol  
 Concen: 4.48 ng  
 RT: 6.73 min Scan# 786  
 Delta R.T. 0.00 min  
 Lab File: BE089398.D  
 Acq: 28 Jan 2015 21:10

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDICC005

Tgt Ion: 112 Resp: 134172  
 Ion Ratio Lower Upper  
 112 100  
 64 56.2 51.7 77.5  
 63 26.5 25.7 38.5




Abundance Ion 112.00 (111.70 to 112.70): BE0  
 Ion 64.00 (63.70 to 64.70): BE0  
 Ion 63.00 (62.70 to 63.70): BE0






#4  
 Phenol-d6  
 Concen: 4.23 ng  
 RT: 8.84 min Scan# 1048  
 Delta R.T. 0.00 min  
 Lab File: BE089398.D  
 Acq: 28 Jan 2015 21:10

Tgt Ion: 99 Resp: 156528  
 Ion Ratio Lower Upper  
 99 100  
 42 17.8 15.3 22.9  
 71 32.1 27.0 40.6



Abundance Ion 99.00 (98.70 to 99.70): BE0  
 Ion 42.00 (41.70 to 42.70): BE0  
 Ion 71.00 (70.70 to 71.70): BE0





#5

Phenol

Concen: 4.51 ng

RT: 8.86 min Scan# 1051

Delta R.T. 0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10

Instrument :

BNA\_E

LabSampleId :

SSTDIC005





Tgt Ion: 94 Resp: 168335

Ion Ratio Lower Upper

94 100

65 148.8 148.6 188.6

66 58.0 46.2 86.2



#6

bis(2-Chloroethyl)ether

Concen: 4.46 ng

RT: 9.06 min Scan# 1073

Delta R.T. 0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10




Tgt Ion: 93 Resp: 124674


Ion Ratio Lower Upper

93 100

63 75.3 58.9 88.3

95 0.0 0.0 0.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 12.50 min Scan# 1563

Delta R.T. 0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10

Instrument :

BNA\_E

LabSampleId :

SSTDIC005


Tgt Ion: 136 Resp: 286326

Ion Ratio Lower Upper

136 100

137 11.0 8.7 13.1




54 8.7 7.4 11.0



Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0



#8

Nitrobenzene-d5

Concen: 4.53 ng

RT: 10.84 min Scan# 1349

Delta R.T. -0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10


Tgt Ion: 82 Resp: 100534

Ion Ratio Lower Upper

82 100

128 41.3 33.6 50.4



54 50.4 37.7 56.5




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Nitrobenzene

Concen: 4.62 ng

RT: 10.89 min Scan# 1359

Delta R.T. 0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10

Instrument :

BNA\_E

LabSampleId :

SSTDICC005


Tgt Ion: 77 Resp: 114639

Ion Ratio Lower Upper

77 100

123 44.5 34.7 52.1




65 13.5 10.6 15.8



Abundance Ion 77.00 (76.70 to 77.70): BE0

Ion 123.00 (122.70 to 123.70): BE0

Ion 65.00 (64.70 to 65.70): BE0



#10

2,4-Dimethylphenol

Concen: 4.11 ng

RT: 11.82 min Scan# 1493

Delta R.T. 0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10


Tgt Ion: 122 Resp: 84516

Ion Ratio Lower Upper

122 100

107 111.3 89.4 134.2



121 54.7 43.6 65.4




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E









#11  
 2,4-Dichlorophenol  
 Concen: 4.59 ng  
 RT: 12.20 min Scan# 1532  
 Delta R.T. -0.01 min  
 Lab File: BE089398.D  
 Acq: 28 Jan 2015 21:10




Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDICC005

Tgt Ion:162 Resp: 66152  
 Ion Ratio Lower Upper  
 162 100  
 164 60.0 44.6 84.6  
 98 45.1 8.7 48.7




#12  
 Hexachlorobutadiene  
 Concen: 4.82 ng  
 RT: 12.89 min Scan# 1603  
 Delta R.T. 0.00 min  
 Lab File: BE089398.D  
 Acq: 28 Jan 2015 21:10

Tgt Ion:225 Resp: 45913  
 Ion Ratio Lower Upper  
 225 100  
 223 63.4 50.2 75.4  
 227 63.4 50.5 75.7











#13  
Acenaphthene-d10  
Concen: 5.00 ng  
RT: 12.21 min Scan# 1533  
Delta R.T. 0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10




Instrument :  
BNA\_E  
LabSampleId :  
SSTDIC005


Tgt Ion:164 Resp: 41601  
Ion Ratio Lower Upper  
164 100  
162 150.7 123.9 185.9  
160 0.0 0.0 0.0



#14  
2,4,6-Tribromophenol  
Concen: 3.74 ng  
RT: 18.58 min Scan# 3037  
Delta R.T. 0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10

Tgt Ion:330 Resp: 10960  
Ion Ratio Lower Upper  
330 100  
332 95.9 77.5 116.3  
141 40.4 32.4 48.6









#15  
2-Fluorobiphenyl  
Concen: 3.73 ng  
RT: 15.13 min Scan# 2427  
Delta R.T. 0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10




Instrument :  
BNA\_E  
LabSampleId :  
SSTDIC005


Tgt Ion:172 Resp: 149366  
Ion Ratio Lower Upper  
172 100  
171 32.1 25.6 38.4  
170 21.3 17.0 25.4



#16  
2,4-Dinitrophenol  
Concen: 5.74 ng  
RT: 16.90 min Scan# 2789  
Delta R.T. 0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10

Tgt Ion:184 Resp: 4779  
Ion Ratio Lower Upper  
184 100  
63 70.3 66.9 100.3  
154 48.9 45.1 67.7









#17  
Phenanthrene-d10  
Concen: 5.00 ng  
RT: 19.98 min Scan# 3241  
Delta R.T. 0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10




Instrument :  
BNA\_E  
LabSampleId :  
SSTDIC005


Tgt Ion:188 Resp: 167125  
Ion Ratio Lower Upper  
188 100  
94 14.6 12.0 18.0  
80 16.1 13.2 19.8



#18  
Hexachlorobenzene  
Concen: 4.46 ng  
RT: 19.23 min Scan# 3130  
Delta R.T. 0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10

Tgt Ion:284 Resp: 37152  
Ion Ratio Lower Upper  
284 100  
142 38.4 30.7 46.1  
249 29.7 23.4 35.2









#19  
 Pentachlorophenol  
 Concen: 7.70 ng  
 RT: 19.65 min Scan# 3192  
 Delta R.T. 0.00 min  
 Lab File: BE089398.D  
 Acq: 28 Jan 2015 21:10




Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDICC005


Tgt Ion: 266 Resp: 11054  
 Ion Ratio Lower Upper  
 266 100  
 268 64.5 52.5 78.7  
 264 61.1 48.8 73.2



#20  
 Chrysene-d12  
 Concen: 5.00 ng  
 RT: 23.84 min Scan# 3874  
 Delta R.T. -0.00 min  
 Lab File: BE089398.D  
 Acq: 28 Jan 2015 21:10

Tgt Ion: 240 Resp: 290825  
 Ion Ratio Lower Upper  
 240 100  
 120 9.8 7.1 10.7  
 236 24.7 19.5 29.3





#21

Benzidine

Concen: 6.10 ng

RT: 22.14 min Scan# 3601

Delta R.T. 0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10

Instrument :

BNA\_E

LabSampleId :

SSTDIC005


Tgt Ion:184 Resp: 139101

Ion Ratio Lower Upper

184 100

185 15.3 11.4 17.0




183 10.1 8.7 13.1



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#22

Terphenyl-d14

Concen: 5.22 ng

RT: 22.48 min Scan# 3676

Delta R.T. -0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10


Tgt Ion:244 Resp: 138545

Ion Ratio Lower Upper

244 100

212 7.2 5.9 8.9



122 14.3 12.8 19.2




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#23

Benzo(a)anthracene

Concen: 6.66 ng

RT: 23.83 min Scan# 3872

Delta R.T. -0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10

Instrument :

BNA\_E

LabSampleId :

SSTDIC005


Tgt Ion:228 Resp: 1347792

Ion Ratio Lower Upper

228 100

226 26.3 21.8 32.6




229 19.0 14.5 21.7



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#24

3,3'-Dichlorobenzidine

Concen: 7.58 ng

RT: 23.81 min Scan# 3870

Delta R.T. -0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10


Tgt Ion:252 Resp: 137172

Ion Ratio Lower Upper

252 100

254 60.7 50.2 75.4



126 13.0 9.8 14.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#25

Chrysene

Concen: 4.78 ng

RT: 23.87 min Scan# 3877

Delta R.T. -0.00 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10

Instrument :

BNA\_E

LabSampleId :

SSTDIC005


Tgt Ion:228 Resp: 1581239

Ion Ratio Lower Upper

228 100

226 28.9 22.7 34.1




229 18.7 15.2 22.8



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#26

Indeno(1,2,3-cd)pyrene

Concen: 4.76 ng

RT: 27.07 min Scan# 4231

Delta R.T. 0.01 min

Lab File: BE089398.D

Acq: 28 Jan 2015 21:10


Tgt Ion:276 Resp: 555936

Ion Ratio Lower Upper

276 100

138 33.1 0.0 0.0#



277 24.6 0.0 0.0#




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#27  
Perylene-d12  
Concen: 5.00 ng  
RT: 25.55 min Scan# 4063  
Delta R.T. -0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10

Instrument :  
BNA\_E  
LabSampleId :  
SSTDIC005

Tgt Ion: 264 Resp: 368965  




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 264 | 100   |       |       |
| 260 | 23.6  | 18.9  | 28.3  |
| 265 | 21.8  | 17.5  | 26.3  |




#28  
Benzo(b)fluoranthene  
Concen: 12.74 ng  
RT: 25.10 min Scan# 4013  
Delta R.T. -0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10

Tgt Ion: 252 Resp: 318479  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 252 | 100   |       |       |
| 253 | 22.4  | 17.6  | 26.4  |
| 125 | 11.5  | 9.8   | 14.6  |











#29  
Benzo(k)fluoranthene  
Concen: 10.23 ng  
RT: 25.13 min Scan# 4016  
Delta R.T. -0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10




Instrument :  
BNA\_E  
LabSampleId :  
SSTDIC005


Tgt Ion: 252 Resp: 580746  
Ion Ratio Lower Upper  
252 100  
253 21.6 17.3 25.9  
125 13.0 10.2 15.4



#30  
Benzo(a)pyrene  
Concen: 11.55 ng  
RT: 25.48 min Scan# 4055  
Delta R.T. -0.00 min  
Lab File: BE089398.D  
Acq: 28 Jan 2015 21:10

Tgt Ion: 252 Resp: 384889  
Ion Ratio Lower Upper  
252 100  
253 21.1 17.5 26.3  
125 15.3 12.2 18.4





#31

Dibenzo(a,h)anthracene

Concen: 15.76 ng

RT: 27.09 min Scan# 4234

Delta R.T. 0.01 min

Lab File: BE089398.D


Acq: 28 Jan 2015 21:10

Instrument :

BNA\_E

LabSampleId :

SSTDIC005




Tgt Ion: 278 Resp: 421559

Ion Ratio Lower Upper

278 100

139 19.5 18.5 27.7

279 23.6 18.6 28.0



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E

