

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE012815\
 Data File : BE089392.D
 Acq On : 28 Jan 2015 17:14
 Operator : TP/IZ
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.1

Quant Time: Jan 29 12:41:14 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 29 11:58:27 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.58	152	129119	5.00	ng	0.00
7) Naphthalene-d8	12.50	136	470387	5.00	ng	0.00
13) Acenaphthene-d10	16.68	164	218523	5.00	ng	0.00
17) Phenanthrene-d10	19.98	188	400455	5.00	ng	0.00
20) Chrysene-d12	23.84	240	485916	5.00	ng	0.00
27) Perylene-d12	25.55	264	544643	5.00	ng	0.00

System Monitoring Compounds						
3) 2-Fluorophenol	6.73	112	4944	0.12	ng	0.00
4) Phenol-d6	8.84	99	6378	0.13	ng	0.00
8) Nitrobenzene-d5	10.84	82	3803	0.11	ng	0.00
14) 2,4,6-Tribromophenol	18.58	330	567	0.00	ng	0.00
15) 2-Fluorobiphenyl	15.13	172	6842	0.00	ng	0.00
22) Terphenyl-d14	22.48	244	6202	0.12	ng	0.00

Target Compounds				Qvalue		
2) n-Nitrosodimethylamine	3.63	42	2351m	0.10	ng	
5) Phenol	8.87	94	6572	0.12	ng	# 58
6) bis(2-Chloroethyl)ether	9.06	93	4791	0.12	ng	# 98
9) Nitrobenzene	10.89	77	4163	0.11	ng	99
10) 2,4-Dimethylphenol	11.82	122	3307	0.12	ng	99
11) 2,4-Dichlorophenol	12.21	162	2554	0.12	ng	96
12) Hexachlorobutadiene	12.89	225	1832	0.12	ng	99
18) Hexachlorobenzene	19.23	284	2049	0.11	ng	98
19) Pentachlorophenol	19.65	266	361	0.10	ng	91
23) Benzo(a)anthracene	23.83	228	50533	0.11	ng	100
24) 3,3'-Dichlorobenzidine	23.81	252	3126	0.09	ng	# 96
25) Chrysene	23.87	228	61805	0.11	ng	99
26) Indeno(1,2,3-cd)pyrene	27.05	276	7857	0.06	ng	# 100
28) Benzo(b)fluoranthene	25.09	252	5767m	0.08	ng	
29) Benzo(k)fluoranthene	25.13	252	13691m	0.08	ng	
30) Benzo(a)pyrene	25.48	252	6451	0.08	ng	# 92
31) Dibenzo(a,h)anthracene	27.08	278	5908	0.07	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE012815\
Data File : BE089392.D
Acq On : 28 Jan 2015 17:14
Operator : TP/IZ
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.1

Quant Time: Jan 29 12:41:14 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 29 11:58:27 2015
Response via : Initial Calibration

