

Data Path : Z:\HPCHEM1\BNA_E\Data\BE012815\
 Data File : BE089400.D
 Acq On : 28 Jan 2015 22:30
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE012815

Quant Time: Jan 29 13:28:21 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 29 13:20:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	67	0.00
2	n-Nitrosodimethylamine	1.000	0.923	7.7	65	0.01
3 S	2-Fluorophenol	1.000	0.946	5.4	68	0.00
4 S	Phenol-d6	1.000	0.937	6.3	67	0.00
5 C	Phenol	1.000	0.942	5.8	66	0.00
6	bis(2-Chloroethyl)ether	1.000	0.960	4.0	68	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	64	0.00
8 S	Nitrobenzene-d5	1.000	0.966	3.4	67	0.00
9	Nitrobenzene	1.000	0.965	3.5	66	0.00
10	2,4-Dimethylphenol	1.000	0.942	5.8	64	0.00
11 C	2,4-Dichlorophenol	1.000	0.928	7.2	63	0.00
12	Hexachlorobutadiene	1.000	0.962	3.8	68	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	59	0.00
14 S	2,4,6-Tribromophenol	1.000	0.991	0.9	60	0.00
15 S	2-Fluorobiphenyl	1.000	0.982	1.8	63	0.00
16 P	2,4-Dinitrophenol	1.000	0.948	5.2	50	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	55	0.00
18	Hexachlorobenzene	1.000	0.954	4.6	57	0.00
19 C	Pentachlorophenol	1.000	0.952	4.8	62	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	64	0.00
21	Benzidine	1.000	0.774	22.6	58	0.00
22 S	Terphenyl-d14	1.000	0.945	5.5	57	0.00
23	Benzo(a)anthracene	1.000	0.891	10.9	60	0.00
24	3,3'-Dichlorobenzidine	1.000	0.851	14.9	59	0.00
25	Chrysene	1.000	0.992	0.8	66	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	0.866	13.4	60	0.00
27 I	Perylene-d12	5.000	5.000	0.0	67	0.00
28	Benzo(b)fluoranthene	1.000	0.762	23.8	56	0.00
29	Benzo(k)fluoranthene	1.000	0.920	8.0	61	0.00
30 C	Benzo(a)pyrene	1.000	0.820	18.0	59	0.00
31	Dibenzo(a,h)anthracene	1.000	0.806	19.4	59	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0