

Data Path : Z:\HPCHEM1\BNA_E\Data\BE012815\
 Data File : BE089400.D
 Acq On : 28 Jan 2015 22:30
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE012815

Quant Time: Jan 29 13:28:21 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 29 13:20:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	67	0.00
2	n-Nitrosodimethylamine	0.822	0.759	7.7	65	0.01
3 S	2-Fluorophenol	1.563	1.479	5.4	68	0.00
4 S	Phenol-d6	1.926	1.806	6.2	67	0.00
5 C	Phenol	2.047	1.928	5.8	66	0.00
6	bis(2-Chloroethyl)ether	1.523	1.462	4.0	68	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	64	0.00
8 S	Nitrobenzene-d5	0.352	0.340	3.4	67	0.00
9	Nitrobenzene	0.398	0.384	3.5	66	0.00
10	2,4-Dimethylphenol	0.303	0.286	5.6	64	0.00
11 C	2,4-Dichlorophenol	0.233	0.217	6.9	63	0.00
12	Hexachlorobutadiene	0.166	0.159	4.2	68	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	59	0.00
14 S	2,4,6-Tribromophenol	0.114	0.112	1.8	60	0.00
15 S	2-Fluorobiphenyl	1.517	1.489	1.8	63	0.00
16 P	2,4-Dinitrophenol	0.034	0.022#	35.3#	50#	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	55	0.00
18	Hexachlorobenzene	0.230	0.220	4.3	57	0.00
19 C	Pentachlorophenol	0.044	0.040	9.1	62	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	64	0.00
21	Benzidine	0.415	0.321	22.7	58	0.00
22 S	Terphenyl-d14	0.554	0.523	5.6	57	0.00
23	Benzo(a)anthracene	4.630	4.126	10.9	60	0.00
24	3,3'-Dichlorobenzidine	0.369	0.304	17.6	59	0.00
25	Chrysene	5.662	5.619	0.8	66	0.00
26	Indeno(1,2,3-cd)pyrene	1.325	0.981	26.0#	60	0.00
27 I	Perylene-d12	1.000	1.000	0.0	67	0.00
28	Benzo(b)fluoranthene	0.659	0.508	22.9	56	0.00
29	Benzo(k)fluoranthene	1.433	1.319	8.0	61	0.00
30 C	Benzo(a)pyrene	0.789	0.618	21.7#	59	0.00
31	Dibenzo(a,h)anthracene	0.805	0.613	23.9	59	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 1