

Data Path : Z:\HPCHEM1\BNA_E\Data\BE012815\
 Data File : BE089407.D
 Acq On : 29 Jan 2015 3:02
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Jan 29 05:35:55 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 29 04:21:19 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	94	0.00
2	n-Nitrosodimethylamine	1.000	0.914	8.6	98	0.02
3 S	2-Fluorophenol	1.000	0.958	4.2	97	0.00
4 S	Phenol-d6	1.000	0.954	4.6	96	0.01
5 C	Phenol	1.000	0.960	4.0	95	0.01
6	bis(2-Chloroethyl)ether	1.000	0.945	5.5	94	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	92	0.00
8 S	Nitrobenzene-d5	1.000	0.947	5.3	94	0.00
9	Nitrobenzene	1.000	0.947	5.3	92	0.00
10	2,4-Dimethylphenol	1.000	0.979	2.1	96	0.00
11 C	2,4-Dichlorophenol	1.000	0.969	3.1	94	0.00
12	Hexachlorobutadiene	1.000	0.913	8.7	92	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	7	-4.47#
14 S	2,4,6-Tribromophenol	1.000	0.373	62.7#	93	0.00
15 S	2-Fluorobiphenyl	1.000	0.372	62.8#	91	0.00
16 P	2,4-Dinitrophenol	1.000	0.963	3.7	63	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	90	0.00
18	Hexachlorobenzene	1.000	0.923	7.7	90	0.00
19 C	Pentachlorophenol	1.000	0.784	21.6#	87	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	94	0.00
21	Benzidine	1.000	0.801	19.9	125	0.00
22 S	Terphenyl-d14	1.000	1.051	-5.1	93	0.00
23	Benzo(a)anthracene	1.000	1.045	-4.5	103	0.00
24	3,3'-Dichlorobenzidine	1.000	1.216	-21.6	127	0.00
25	Chrysene	1.000	0.935	6.5	92	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	1.159	-15.9	137	0.00
27 I	Perylene-d12	5.000	5.000	0.0	102	0.00
28	Benzo(b)fluoranthene	1.000	1.299	-29.9#	143	0.00
29	Benzo(k)fluoranthene	1.000	0.930	7.0	105	0.00
30 C	Benzo(a)pyrene	1.000	1.219	-21.9#	135	0.00
31	Dibenzo(a,h)anthracene	1.000	1.250	-25.0	149	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 2