

Data Path : Z:\HPCHEM1\BNA_E\Data\BE012815\
 Data File : BE089407.D
 Acq On : 29 Jan 2015 3:02
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Jan 29 05:35:55 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 29 04:21:19 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	n-Nitrosodimethylamine	0.822	0.817	0.6	98	0.02
3 S	2-Fluorophenol	1.563	1.497	4.2	97	0.00
4 S	Phenol-d6	1.926	1.837	4.6	96	0.01
5 C	Phenol	2.047	1.964	4.1	95	0.01
6	bis(2-Chloroethyl)ether	1.523	1.439	5.5	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.352	0.334	5.1	94	0.00
9	Nitrobenzene	0.398	0.377	5.3	92	0.00
10	2,4-Dimethylphenol	0.303	0.297	2.0	96	0.00
11 C	2,4-Dichlorophenol	0.233	0.226	3.0	94	0.00
12	Hexachlorobutadiene	0.166	0.151	9.0	92	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	7#	-4.47#
14 S	2,4,6-Tribromophenol	0.114	1.571	-1278.1#	93	0.00
15 S	2-Fluorobiphenyl	1.517	19.672	-1196.8#	91	0.00
16 P	2,4-Dinitrophenol	0.034	0.251	-638.2#	63	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
18	Hexachlorobenzene	0.230	0.212	7.8	90	0.00
19 C	Pentachlorophenol	0.044	0.035	20.5#	87	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
21	Benzidine	0.415	0.473	-14.0	125	0.00
22 S	Terphenyl-d14	0.554	0.582	-5.1	93	0.00
23	Benzo(a)anthracene	4.630	4.837	-4.5	103	0.00
24	3,3'-Dichlorobenzidine	0.369	0.448	-21.4	127	0.00
25	Chrysene	5.662	5.293	6.5	92	0.00
26	Indeno(1,2,3-cd)pyrene	1.325	1.530	-15.5	137	0.00
27 I	Perylene-d12	1.000	1.000	0.0	102	0.00
28	Benzo(b)fluoranthene	0.659	0.853	-29.4#	143	0.00
29	Benzo(k)fluoranthene	1.433	1.481	-3.3	105	0.00
30 C	Benzo(a)pyrene	0.789	0.940	-19.1	135	0.00
31	Dibenzo(a,h)anthracene	0.805	1.025	-27.3#	149	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1