

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE012915\
 Data File : BE089414.D
 Acq On : 29 Jan 2015 13:58
 Operator : TP/IZ
 Sample : SSTDICC002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC002

Manual Integrations
 APPROVED

apatel
 1/30/2015 5:55:37 PM

Quant Time: Jan 30 03:01:40 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 30 02:34:28 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.58	152	119016	5.00	ng	0.00
7) Naphthalene-d8	12.50	136	438711	5.00	ng	0.00
13) Acenaphthene-d10	16.68	164	166556	5.00	ng	0.00
17) Phenanthrene-d10	19.99	188	331190	5.00	ng	0.00
20) Chrysene-d12	23.84	240	451988	5.00	ng	0.00
27) Perylene-d12	25.55	264	463111	5.00	ng	0.00
System Monitoring Compounds						
3) 2-Fluorophenol	6.73	112	92519	2.42	ng	0.00
4) Phenol-d6	8.85	99	109117	2.34	ng	0.00
8) Nitrobenzene-d5	10.84	82	75215	2.38	ng	0.00
14) 2,4,6-Tribromophenol	18.58	330	10081	2.60	ng	0.00
15) 2-Fluorobiphenyl	15.13	172	98192	1.95	ng	0.00
22) Terphenyl-d14	22.48	244	116752	2.34	ng	0.00
Target Compounds						Qvalue
2) n-Nitrosodimethylamine	3.57	42	54783	2.70	ng	# 84
5) Phenol	8.87	94	120326	2.42	ng	99
6) bis(2-Chloroethyl)ether	9.06	93	89431	2.42	ng	# 99
9) Nitrobenzene	10.88	77	86524	2.43	ng	100
10) 2,4-Dimethylphenol	11.82	122	60706	2.26	ng	99
11) 2,4-Dichlorophenol	12.21	162	42856	2.09	ng	97
12) Hexachlorobutadiene	12.89	225	27984	1.93	ng	99
16) 2,4-Dinitrophenol	16.89	184	2536	2.25	ng	95
18) Hexachlorobenzene	19.23	284	33607	2.18	ng	94
19) Pentachlorophenol	19.65	266	7338	2.48	ng	95
21) Benzidine	22.14	184	66864	1.83	ng	98
23) Benzo(a)anthracene	23.83	228	721409	1.76	ng	99
24) 3,3'-Dichlorobenzidine	23.81	252	61543	1.88	ng	98
25) Chrysene	23.87	228	991110	1.96	ng	98
26) Indeno(1,2,3-cd)pyrene	27.08	276	187338	1.60	ng	# 100
28) Benzo(b)fluoranthene	25.10	252	106604m	1.82	ng	
29) Benzo(k)fluoranthene	25.13	252	320489m	2.41	ng	
30) Benzo(a)pyrene	25.48	252	133813	1.88	ng	99
31) Dibenzo(a,h)anthracene	27.09	278	138856	1.91	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE012915\
Data File : BE089414.D
Acq On : 29 Jan 2015 13:58
Operator : TP/IZ
Sample : SSTDICC002
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC002

Manual Integrations
APPROVED

apatel
1/30/2015 5:55:37 PM

Quant Time: Jan 30 03:01:40 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 30 02:34:28 2015
Response via : Initial Calibration

