

Data Path : Z:\HPCHEM1\BNA_E\Data\BE012915\
 Data File : BE089423.D
 Acq On : 29 Jan 2015 21:02
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 30 07:27:24 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 30 03:48:17 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	96	0.00
2	n-Nitrosodimethylamine	1.000	1.018	-1.8	94	0.00
3 S	2-Fluorophenol	1.000	0.927	7.3	95	0.00
4 S	Phenol-d6	1.000	0.939	6.1	96	0.00
5 C	Phenol	1.000	0.932	6.8	96	0.00
6	bis(2-Chloroethyl)ether	1.000	0.925	7.5	95	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	98	0.00
8 S	Nitrobenzene-d5	1.000	0.933	6.7	96	0.00
9	Nitrobenzene	1.000	0.951	4.9	97	0.00
10	2,4-Dimethylphenol	1.000	0.949	5.1	98	0.00
11 C	2,4-Dichlorophenol	1.000	0.954	4.6	97	0.00
12	Hexachlorobutadiene	1.000	0.933	6.7	98	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	1.000	0.914	8.6	98	0.00
15 S	2-Fluorobiphenyl	1.000	0.943	5.7	101	0.00
16 P	2,4-Dinitrophenol	1.000	0.806	19.4	124	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	105	0.00
18	Hexachlorobenzene	1.000	0.951	4.9	106	0.00
19 C	Pentachlorophenol	1.000	0.684	31.6#	86	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	103	0.00
21	Benzidine	1.000	0.801	19.9	126	0.00
22 S	Terphenyl-d14	1.000	1.077	-7.7	111	0.00
23	Benzo(a)anthracene	1.000	1.102	-10.2	118	0.00
24	3,3'-Dichlorobenzidine	1.000	1.173	-17.3	132	0.00
25	Chrysene	1.000	0.923	7.7	101	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	1.273	-27.3#	147	0.00
27 I	Perylene-d12	5.000	5.000	0.0	102	0.00
28	Benzo(b)fluoranthene	1.000	1.371	-37.1#	170	0.00
29	Benzo(k)fluoranthene	1.000	1.075	-7.5	99	0.00
30 C	Benzo(a)pyrene	1.000	1.237	-23.7#	142	0.00
31	Dibenzo(a,h)anthracene	1.000	1.392	-39.2#	160	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 2