

Data Path : Z:\HPCHEM1\BNA_E\Data\BE012915\
 Data File : BE089423.D
 Acq On : 29 Jan 2015 21:02
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 30 07:27:24 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 30 03:48:17 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	n-Nitrosodimethylamine	0.978	0.995	-1.7	94	0.00
3 S	2-Fluorophenol	1.881	1.744	7.3	95	0.00
4 S	Phenol-d6	2.235	2.098	6.1	96	0.00
5 C	Phenol	2.452	2.285	6.8	96	0.00
6	bis(2-Chloroethyl)ether	1.821	1.685	7.5	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.406	0.379	6.7	96	0.00
9	Nitrobenzene	0.464	0.442	4.7	97	0.00
10	2,4-Dimethylphenol	0.329	0.313	4.9	98	0.00
11 C	2,4-Dichlorophenol	0.229	0.218	4.8	97	0.00
12	Hexachlorobutadiene	0.154	0.144	6.5	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.138	0.126	8.7	98	0.00
15 S	2-Fluorobiphenyl	1.395	1.316	5.7	101	0.00
16 P	2,4-Dinitrophenol	0.037	0.030#	18.9	124	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
18	Hexachlorobenzene	0.245	0.233	4.9	106	0.00
19 C	Pentachlorophenol	0.046	0.031	32.6#	86	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
21	Benzidine	0.437	0.350	19.9	126	0.00
22 S	Terphenyl-d14	0.578	0.623	-7.8	111	0.00
23	Benzo(a)anthracene	3.805	4.195	-10.2	118	0.00
24	3,3'-Dichlorobenzidine	0.299	0.350	-17.1	132	0.00
25	Chrysene	5.201	4.800	7.7	101	0.00
26	Indeno(1,2,3-cd)pyrene	0.897	1.143	-27.4#	147	0.00
27 I	Perylene-d12	1.000	1.000	0.0	102	0.00
28	Benzo(b)fluoranthene	0.493	0.676	-37.1#	170#	0.00
29	Benzo(k)fluoranthene	1.302	1.399	-7.5	99	0.00
30 C	Benzo(a)pyrene	0.610	0.755	-23.8#	142	0.00
31	Dibenzo(a,h)anthracene	0.608	0.846	-39.1#	160#	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 2