

Data Path : Z:\HPCHEM1\BNA_E\Data\BE012915\
 Data File : BE089425.D
 Acq On : 29 Jan 2015 23:01
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 30 17:16:30 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 30 17:07:09 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	n-Nitrosodimethylamine	0.978	0.970	0.8	86	0.00
3 S	2-Fluorophenol	1.881	1.718	8.7	88	0.00
4 S	Phenol-d6	2.235	2.090	6.5	91	0.00
5 C	Phenol	2.452	2.300	6.2	91	0.00
6	bis(2-Chloroethyl)ether	1.821	1.702	6.5	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.406	0.387	4.7	94	0.00
9	Nitrobenzene	0.464	0.445	4.1	95	0.00
10	2,4-Dimethylphenol	0.329	0.313	4.9	95	0.00
11 C	2,4-Dichlorophenol	0.229	0.223	2.6	96	0.00
12	Hexachlorobutadiene	0.154	0.144	6.5	94	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.138	0.138	0.0	107	0.00
15 S	2-Fluorobiphenyl	1.395	1.312	5.9	100	0.00
16 P	2,4-Dinitrophenol	0.037	0.037#	0.0	155#	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
18	Hexachlorobenzene	0.245	0.236	3.7	105	0.00
19 C	Pentachlorophenol	0.046	0.037	19.6	100	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
21	Benzidine	0.437	0.485	-11.0	157#	0.00
22 S	Terphenyl-d14	0.578	0.632	-9.3	102	0.00
23	Benzo(a)anthracene	3.805	3.992	-4.9	102	0.00
24	3,3'-Dichlorobenzidine	0.299	0.330	-10.4	112	0.00
25	Chrysene	5.201	4.817	7.4	92	0.00
26	Indeno(1,2,3-cd)pyrene	0.897	1.014	-13.0	118	0.00
27 I	Perylene-d12	1.000	1.000	0.0	92	0.00
28	Benzo(b)fluoranthene	0.493	2.042	-314.2#	459#	0.03
29	Benzo(k)fluoranthene	1.302	2.059	-58.1#	130	0.00
30 C	Benzo(a)pyrene	0.610	0.673	-10.3	113	0.00
31	Dibenzo(a,h)anthracene	0.608	0.742	-22.0	126	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 0