

Data Path : Z:\HPCHEM1\BNA_E\Data\BE012915\
 Data File : BE089431.D
 Acq On : 30 Jan 2015 3:36
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 30 07:52:14 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 30 03:48:17 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	94	0.00
2	n-Nitrosodimethylamine	1.000	0.993	0.7	89	0.00
3 S	2-Fluorophenol	1.000	0.923	7.7	92	0.02
4 S	Phenol-d6	1.000	0.941	5.9	94	0.03
5 C	Phenol	1.000	0.932	6.8	94	0.03
6	bis(2-Chloroethyl)ether	1.000	0.882	11.8	89	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	96	0.00
8 S	Nitrobenzene-d5	1.000	0.919	8.1	92	0.00
9	Nitrobenzene	1.000	0.937	6.3	93	0.00
10	2,4-Dimethylphenol	1.000	0.991	0.9	100	0.02
11 C	2,4-Dichlorophenol	1.000	1.014	-1.4	100	0.00
12	Hexachlorobutadiene	1.000	0.931	6.9	95	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	1.000	1.079	-7.9	118	0.01
15 S	2-Fluorobiphenyl	1.000	0.929	7.1	102	0.00
16 P	2,4-Dinitrophenol	1.000	0.878	12.2	138	0.01
17 I	Phenanthrene-d10	5.000	5.000	0.0	114	0.00
18	Hexachlorobenzene	1.000	0.944	5.6	114	0.00
19 C	Pentachlorophenol	1.000	0.689	31.1#	93	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	108	0.00
21	Benzidine	1.000	0.950	5.0	157	0.00
22 S	Terphenyl-d14	1.000	1.150	-15.0	124	0.00
23	Benzo(a)anthracene	1.000	1.202	-20.2	135	0.00
24	3,3'-Dichlorobenzidine	1.000	1.406	-40.6#	166	0.00
25	Chrysene	1.000	0.909	9.1	105	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	1.579	-57.9#	192	0.01
27 I	Perylene-d12	5.000	5.000	0.0	108	0.00
28	Benzo(b)fluoranthene	1.000	1.973	-97.3#	257	0.00
29	Benzo(k)fluoranthene	1.000	1.027	-2.7	99	0.00
30 C	Benzo(a)pyrene	1.000	1.641	-64.1#	198	0.00
31	Dibenzo(a,h)anthracene	1.000	1.843	-84.3#	223	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 2