

Data Path : Z:\HPCHEM1\BNA_E\Data\BE012915\
 Data File : BE089431.D
 Acq On : 30 Jan 2015 3:36
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 30 07:52:14 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE012915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 30 03:48:17 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	n-Nitrosodimethylamine	0.978	0.971	0.7	89	0.00
3 S	2-Fluorophenol	1.881	1.737	7.7	92	0.02
4 S	Phenol-d6	2.235	2.103	5.9	94	0.03
5 C	Phenol	2.452	2.285	6.8	94	0.03
6	bis(2-Chloroethyl)ether	1.821	1.605	11.9	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.406	0.374	7.9	92	0.00
9	Nitrobenzene	0.464	0.435	6.3	93	0.00
10	2,4-Dimethylphenol	0.329	0.327	0.6	100	0.02
11 C	2,4-Dichlorophenol	0.229	0.232	-1.3	100	0.00
12	Hexachlorobutadiene	0.154	0.144	6.5	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.138	0.149	-8.0	118	0.01
15 S	2-Fluorobiphenyl	1.395	1.295	7.2	102	0.00
16 P	2,4-Dinitrophenol	0.037	0.032#	13.5	138	0.01
17 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
18	Hexachlorobenzene	0.245	0.231	5.7	114	0.00
19 C	Pentachlorophenol	0.046	0.031	32.6#	93	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
21	Benzidine	0.437	0.416	4.8	157#	0.00
22 S	Terphenyl-d14	0.578	0.665	-15.1	124	0.00
23	Benzo(a)anthracene	3.805	4.574	-20.2	135	0.00
24	3,3'-Dichlorobenzidine	0.299	0.420	-40.5#	166#	0.00
25	Chrysene	5.201	4.728	9.1	105	0.00
26	Indeno(1,2,3-cd)pyrene	0.897	1.417	-58.0#	192#	0.01
27 I	Perylene-d12	1.000	1.000	0.0	108	0.00
28	Benzo(b)fluoranthene	0.493	0.972	-97.2#	257#	0.00
29	Benzo(k)fluoranthene	1.302	1.337	-2.7	99	0.00
30 C	Benzo(a)pyrene	0.610	1.001	-64.1#	198#	0.00
31	Dibenzo(a,h)anthracene	0.608	1.121	-84.4#	223#	0.01

(#) = Out of Range

SPCC's out = 1 CCC's out = 2