

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
 Data File : BE089433.D
 Acq On : 30 Jan 2015 5:32
 Operator : TP/IZ
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.1

Quant Time: Jan 31 05:08:17 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 04:33:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.59	152	102343	5.00	ng	0.00
2) Naphthalene-d8	12.50	136	414477	5.00	ng	0.00
6) Acenaphthene-d10	16.69	164	187884	5.00	ng	0.00
11) Phenanthrene-d10	19.99	188	385575	5.00	ng	0.00
16) Chrysene-d12	23.84	240	332855	5.00	ng	0.00
22) Perylene-d12	25.56	264	314017	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.85	82	3160	0.15	ng	0.00
7) 2-Fluorobiphenyl	15.14	172	5572	0.13	ng	0.00
18) Terphenyl-d14	22.49	244	6232	0.14	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	12.55	128	11174	0.14	ng	98
5) 2-Methylnaphthalene	14.21	142	6625	0.13	ng	94
8) Acenaphthylene	16.34	152	40599	0.15	ng	99
9) Acenaphthene	16.76	154	6319	0.15	ng	95
10) Fluorene	18.02	166	7120	0.16	ng	100
13) Phenanthrene	20.03	178	47413	0.14	ng	99
14) Anthracene	20.12	178	40832	0.14	ng	100
15) Fluoranthene	21.91	202	10428	0.16	ng	100
17) Pyrene	22.23	202	11535	0.14	ng	100
19) Benzo(a)anthracene	23.83	228	35497	0.16	ng	96
20) Chrysene	23.88	228	39092	0.14	ng	93
21) Indeno(1,2,3-cd)pyrene	27.06	276	22951	0.15	ng	93
23) Benzo(b)fluoranthene	25.10	252	3515	0.10	ng	# 85
24) Benzo(k)fluoranthene	25.13	252	7650	0.12	ng	# 90
25) Benzo(a)pyrene	25.49	252	4255	0.11	ng	# 81
26) Dibenzo(a,h)anthracene	27.09	278	3850	0.12	ng	# 87
27) Benzo(g,h,i)perylene	27.51	276	25561	0.14	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
Data File : BE089433.D
Acq On : 30 Jan 2015 5:32
Operator : TP/IZ
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.1

Quant Time: Jan 31 05:08:17 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 31 04:33:09 2015
Response via : Initial Calibration

