

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
 Data File : BE089435.D
 Acq On : 30 Jan 2015 6:50
 Operator : TP/IZ
 Sample : SSTDICC0.5
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.5

Quant Time: Jan 31 04:43:54 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 04:33:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.59	152	96046	5.00	ng	0.00
2) Naphthalene-d8	12.50	136	394387	5.00	ng	0.00
6) Acenaphthene-d10	16.69	164	175804	5.00	ng	-0.01
11) Phenanthrene-d10	19.99	188	316674	5.00	ng	0.00
16) Chrysene-d12	23.85	240	267091	5.00	ng	0.00
22) Perylene-d12	25.56	264	259451	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.85	82	12223	0.60	ng	0.00
7) 2-Fluorobiphenyl	15.14	172	23059	0.58	ng	0.00
18) Terphenyl-d14	22.49	244	19249	0.56	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	12.54	128	46387	0.61	ng	99
5) 2-Methylnaphthalene	14.21	142	27173	0.58	ng	98
8) Acenaphthylene	16.34	152	152439	0.61	ng	100
9) Acenaphthene	16.76	154	23234	0.59	ng	98
10) Fluorene	18.02	166	25492	0.62	ng	100
12) Pentachlorophenol	19.66	266	1653	0.43	ng	95
13) Phenanthrene	20.03	178	153453	0.57	ng	100
14) Anthracene	20.12	178	136368	0.55	ng	100
15) Fluoranthene	21.91	202	32676	0.61	ng	100
17) Pyrene	22.23	202	34222	0.51	ng	100
19) Benzo(a)anthracene	23.83	228	103534	0.57	ng	99
20) Chrysene	23.87	228	128140	0.57	ng	99
21) Indeno(1,2,3-cd)pyrene	27.06	276	101329	0.81	ng	96
23) Benzo(b)fluoranthene	25.10	252	14348	0.47	ng	98
24) Benzo(k)fluoranthene	25.13	252	34029	0.65	ng	98
25) Benzo(a)pyrene	25.49	252	17120	0.56	ng	# 96
26) Dibenzo(a,h)anthracene	27.09	278	18264	0.71	ng	96
27) Benzo(g,h,i)perylene	27.51	276	95660	0.64	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
Data File : BE089435.D
Acq On : 30 Jan 2015 6:50
Operator : TP/IZ
Sample : SSTDICC0.5
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.5

Quant Time: Jan 31 04:43:54 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 31 04:33:09 2015
Response via : Initial Calibration

