

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
 Data File : BE089436.D
 Acq On : 30 Jan 2015 7:29
 Operator : TP/IZ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.8

Quant Time: Jan 31 04:44:19 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 04:33:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.59	152	88815	5.00	ng	0.00
2) Naphthalene-d8	12.50	136	368331	5.00	ng	0.00
6) Acenaphthene-d10	16.69	164	159330	5.00	ng	-0.01
11) Phenanthrene-d10	19.99	188	267665	5.00	ng	0.00
16) Chrysene-d12	23.84	240	226336	5.00	ng	0.00
22) Perylene-d12	25.55	264	221205	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.85	82	17840	0.94	ng	0.00
7) 2-Fluorobiphenyl	15.13	172	34515	0.95	ng	0.00
18) Terphenyl-d14	22.48	244	26472	0.90	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	12.54	128	68921	0.98	ng	100
5) 2-Methylnaphthalene	14.21	142	40778	0.93	ng	99
8) Acenaphthylene	16.33	152	227631	1.01	ng	100
9) Acenaphthene	16.76	154	34072	0.96	ng	97
10) Fluorene	18.01	166	37415	1.01	ng	98
12) Pentachlorophenol	19.66	266	767	0.24	ng	96
13) Phenanthrene	20.03	178	211432	0.92	ng	100
14) Anthracene	20.12	178	190117	0.91	ng	100
15) Fluoranthene	21.91	202	45007	0.99	ng	99
17) Pyrene	22.23	202	46702	0.82	ng	99
19) Benzo(a)anthracene	23.83	228	147282	0.96	ng	98
20) Chrysene	23.87	228	175113	0.91	ng	98
21) Indeno(1,2,3-cd)pyrene	27.06	276	150332	1.41	ng	99
23) Benzo(b)fluoranthene	25.10	252	22011	0.85	ng	99
24) Benzo(k)fluoranthene	25.13	252	49757	1.11	ng	98
25) Benzo(a)pyrene	25.49	252	27015	1.03	ng	97
26) Dibenzo(a,h)anthracene	27.09	278	30095	1.38	ng	99
27) Benzo(g,h,i)perylene	27.51	276	144571	1.13	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
Data File : BE089436.D
Acq On : 30 Jan 2015 7:29
Operator : TP/IZ
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.8

Quant Time: Jan 31 04:44:19 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 31 04:33:09 2015
Response via : Initial Calibration

