

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
 Data File : BE089439.D
 Acq On : 30 Jan 2015 9:28
 Operator : TP/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC005

Quant Time: Jan 31 04:45:33 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 04:33:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.59	152	80285	5.00	ng	0.00
2) Naphthalene-d8	12.50	136	343300	5.00	ng	0.00
6) Acenaphthene-d10	16.69	164	165600	5.00	ng	-0.01
11) Phenanthrene-d10	19.99	188	278799	5.00	ng	0.00
16) Chrysene-d12	23.85	240	247574	5.00	ng	0.00
22) Perylene-d12	25.55	264	238050	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.85	82	112245	6.33	ng	0.00
7) 2-Fluorobiphenyl	15.14	172	198975	5.29	ng	0.00
18) Terphenyl-d14	22.48	244	160619	5.01	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	12.55	128	372701	5.66	ng	99
5) 2-Methylnaphthalene	14.21	142	233297	5.68	ng	97
8) Acenaphthylene	16.33	152	1357667	5.81	ng	100
9) Acenaphthene	16.76	154	198304	5.36	ng	100
10) Fluorene	18.01	166	219812	5.68	ng	99
12) Pentachlorophenol	19.66	266	9916	2.94	ng	94
13) Phenanthrene	20.03	178	1204154	5.05	ng	100
14) Anthracene	20.12	178	1160392	5.32	ng	100
15) Fluoranthene	21.91	202	269332	5.68	ng	100
17) Pyrene	22.23	202	281263	4.54	ng	100
19) Benzo(a)anthracene	23.83	228	950963	5.69	ng	96
20) Chrysene	23.87	228	1008144	4.81	ng	99
21) Indeno(1,2,3-cd)pyrene	27.06	276	1101712	9.46	ng	100
23) Benzo(b)fluoranthene	25.10	252	197911	7.11	ng	99
24) Benzo(k)fluoranthene	25.13	252	289939	6.01	ng	99
25) Benzo(a)pyrene	25.48	252	204427	7.23	ng	100
26) Dibenzo(a,h)anthracene	27.10	278	230960	9.82	ng	97
27) Benzo(g,h,i)perylene	27.51	276	939827	6.82	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
Data File : BE089439.D
Acq On : 30 Jan 2015 9:28
Operator : TP/IZ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC005

Quant Time: Jan 31 04:45:33 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 31 04:33:09 2015
Response via : Initial Calibration

