

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
 Data File : BE089442.D
 Acq On : 30 Jan 2015 17:22
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE123015

Quant Time: Jan 31 06:18:33 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 05:05:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.59	152	90258	5.00	ng	0.00
2) Naphthalene-d8	12.50	136	357275	5.00	ng	0.00
6) Acenaphthene-d10	16.69	164	150452	5.00	ng	-0.01
11) Phenanthrene-d10	19.99	188	260761	5.00	ng	0.00
16) Chrysene-d12	23.85	240	214742	5.00	ng	0.00
22) Perylene-d12	25.56	264	193833	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.85	82	25923	1.10	ng	0.00
7) 2-Fluorobiphenyl	15.14	172	40882	1.03	ng	0.00
18) Terphenyl-d14	22.48	244	32416	1.01	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	12.54	128	83653	1.00	ng	99
5) 2-Methylnaphthalene	14.21	142	48766	0.96	ng	98
8) Acenaphthylene	16.33	152	274820	1.02	ng	100
9) Acenaphthene	16.76	154	41422	1.02	ng	99
10) Fluorene	18.01	166	44796	0.99	ng	99
12) Pentachlorophenol	19.66	266	49177	13.41	ng	94
13) Phenanthrene	20.03	178	263179	1.03	ng	99
14) Anthracene	20.12	178	236273	1.01	ng	99
15) Fluoranthene	21.92	202	54473	0.97	ng	100
17) Pyrene	22.23	202	61194	1.07	ng	100
19) Benzo(a)anthracene	23.83	228	171035	0.95	ng	99
20) Chrysene	23.87	228	212717	1.04	ng	98
21) Indeno(1,2,3-cd)pyrene	27.06	276	148335	0.83	ng	96
23) Benzo(b)fluoranthene	25.10	252	23397	0.84	ng	98
24) Benzo(k)fluoranthene	25.13	252	58060	1.18	ng	98
25) Benzo(a)pyrene	25.49	252	27185	0.88	ng	99
26) Dibenzo(a,h)anthracene	27.10	278	28122	0.82	ng	98
27) Benzo(g,h,i)perylene	27.51	276	141672	0.91	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
Data File : BE089442.D
Acq On : 30 Jan 2015 17:22
Operator : TP/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE123015

Quant Time: Jan 31 06:18:33 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 31 05:05:18 2015
Response via : Initial Calibration

