

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
 Data File : BE089443.D
 Acq On : 30 Jan 2015 18:01
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 02 11:24:14 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 05:05:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.59	152	86802	5.00	ng	0.00
2) Naphthalene-d8	12.50	136	350538	5.00	ng	0.00
6) Acenaphthene-d10	16.69	164	152472	5.00	ng	-0.01
11) Phenanthrene-d10	19.99	188	275675	5.00	ng	0.00
16) Chrysene-d12	23.84	240	243669	5.00	ng	0.00
22) Perylene-d12	25.55	264	218554	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.85	82	24848	1.07	ng	0.00
7) 2-Fluorobiphenyl	15.13	172	41233	1.03	ng	0.00
18) Terphenyl-d14	22.48	244	35270	0.97	ng	0.00
Target Compounds						Qvalue
4) Naphthalene	12.55	128	81417	0.99	ng	99
5) 2-Methylnaphthalene	14.21	142	48510	0.98	ng	99
8) Acenaphthylene	16.34	152	276648	1.01	ng	100
9) Acenaphthene	16.76	154	42214	1.02	ng	100
10) Fluorene	18.02	166	45819	1.00	ng	99
12) Pentachlorophenol	19.66	266	6776	3.96	ng	93
13) Phenanthrene	20.03	178	275906	1.02	ng	100
14) Anthracene	20.12	178	249765	1.01	ng	99
15) Fluoranthene	21.91	202	59847	1.01	ng	99
17) Pyrene	22.23	202	65404	1.01	ng	100
19) Benzo(a)anthracene	23.83	228	189619	0.92	ng	100
20) Chrysene	23.87	228	238480	1.03	ng	98
21) Indeno(1,2,3-cd)pyrene	27.06	276	173995m	0.86	ng	
23) Benzo(b)fluoranthene	25.10	252	24780	0.80	ng	100
24) Benzo(k)fluoranthene	25.13	252	61322	1.10	ng	99
25) Benzo(a)pyrene	25.48	252	29734	0.85	ng	99
26) Dibenzo(a,h)anthracene	27.09	278	31340	0.81	ng	100
27) Benzo(g,h,i)perylene	27.51	276	157152	0.90	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
Data File : BE089443.D
Acq On : 30 Jan 2015 18:01
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Feb 02 11:24:14 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 31 05:05:18 2015
Response via : Initial Calibration

