

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
 Data File : BE089449.D
 Acq On : 30 Jan 2015 21:57
 Operator : TP/IZ
 Sample : G1121-13
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RBOE-TW1

Quant Time: Feb 02 00:23:08 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 05:05:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.59	152	76537	5.00	ng	0.00
2) Naphthalene-d8	12.50	136	314250	5.00	ng	0.00
6) Acenaphthene-d10	16.69	164	155782	5.00	ng	-0.01
11) Phenanthrene-d10	19.99	188	259614	5.00	ng	0.00
16) Chrysene-d12	23.84	240	227028	5.00	ng	0.00
22) Perylene-d12	25.55	264	209661	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) Nitrobenzene-d5	10.85	82	135225	6.50	ng	0.00
7) 2-Fluorobiphenyl	15.13	172	231487	5.64	ng	0.00
18) Terphenyl-d14	22.48	244	244113	7.18	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Naphthalene	12.54	128	2571	0.03	ng	# 87
5) 2-Methylnaphthalene	14.20	142	1484	0.03	ng	# 94
8) Acenaphthylene	16.34	152	8721	0.03	ng	97
9) Acenaphthene	16.76	154	1714	0.04	ng	# 89
10) Fluorene	18.02	166	1845	0.04	ng	# 93
12) Pentachlorophenol	19.66	266	1057	0.69	ng	93
13) Phenanthrene	20.03	178	15264	0.06	ng	99
14) Anthracene	20.12	178	9835	0.04	ng	97
15) Fluoranthene	21.91	202	2523	0.05	ng	99
17) Pyrene	22.23	202	2724	0.05	ng	98
19) Benzo(a)anthracene	23.83	228	9585	0.05	ng	95
20) Chrysene	23.87	228	10330	0.05	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
Data File : BE089449.D
Acq On : 30 Jan 2015 21:57
Operator : TP/IZ
Sample : G1121-13
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RBOE-TW1

Quant Time: Feb 02 00:23:08 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 31 05:05:18 2015
Response via : Initial Calibration

