

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
 Data File : BE089453.D
 Acq On : 31 Jan 2015 1:11
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 02 00:59:45 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 05:05:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.59	152	92620	5.00	ng	0.00
2) Naphthalene-d8	12.50	136	391925	5.00	ng	0.00
6) Acenaphthene-d10	16.69	164	179257	5.00	ng	-0.01
11) Phenanthrene-d10	19.99	188	335569	5.00	ng	0.00
16) Chrysene-d12	23.84	240	294184	5.00	ng	0.00
22) Perylene-d12	25.55	264	266560	5.00	ng	0.00

System Monitoring Compounds						
3) Nitrobenzene-d5	10.85	82	110081	4.24	ng	0.00
7) 2-Fluorobiphenyl	15.13	172	235540	4.99	ng	0.00
18) Terphenyl-d14	22.48	244	195931	4.45	ng	0.00

Target Compounds					Qvalue	
4) Naphthalene	12.55	128	78319	0.85	ng	100
5) 2-Methylnaphthalene	14.21	142	46664	0.84	ng	100
8) Acenaphthylene	16.34	152	264839	0.82	ng	100
9) Acenaphthene	16.76	154	39672	0.82	ng	99
10) Fluorene	18.02	166	44168	0.82	ng	99
12) Pentachlorophenol	19.66	266	3864	2.24	ng	88
13) Phenanthrene	20.03	178	267932	0.81	ng	99
14) Anthracene	20.12	178	251691	0.84	ng	99
15) Fluoranthene	21.91	202	62819	0.87	ng	99
17) Pyrene	22.22	202	65615	0.84	ng	100
19) Benzo(a)anthracene	23.83	228	202314	0.82	ng	98
20) Chrysene	23.87	228	230668	0.83	ng	98
21) Indeno(1,2,3-cd)pyrene	27.06	276	174950	0.72	ng	100
23) Benzo(b)fluoranthene	25.10	252	30309	0.80	ng	98
24) Benzo(k)fluoranthene	25.13	252	62180	0.92	ng	99
25) Benzo(a)pyrene	25.48	252	32377	0.78	ng	98
26) Dibenzo(a,h)anthracene	27.09	278	35205	0.76	ng	99
27) Benzo(g,h,i)perylene	27.51	276	159776	0.75	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE013015\
Data File : BE089453.D
Acq On : 31 Jan 2015 1:11
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Feb 02 00:59:45 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 31 05:05:18 2015
Response via : Initial Calibration

