

Data Path : Z:\HPCHEM1\BNA_E\Data\BE013015\
 Data File : BE089453.D
 Acq On : 31 Jan 2015 1:11
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 02 00:59:45 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 05:05:18 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	68	0.00
3 S	Nitrobenzene-d5	0.331	1.404	-324.2#	304#	0.00
4	Naphthalene	1.174	0.999	14.9	60	0.00
5	2-Methylnaphthalene	0.709	0.595	16.1	57	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	66	-0.01
7 S	2-Fluorobiphenyl	1.317	6.570	-398.9#	337#	0.00
8	Acenaphthylene	8.993	7.387	17.9	55	0.00
9	Acenaphthene	1.351	1.107	18.1	56	0.00
10	Fluorene	1.502	1.232	18.0	55	0.00
11 I	Phenanthrene-d10	1.000	1.000	0.0	66	0.00
12	Pentachlorophenol	0.035	0.058	-65.7#	145	0.00
13	Phenanthrene	4.913	3.992	18.7	55	0.00
14	Anthracene	4.487	3.750	16.4	55	0.00
15	Fluoranthene	1.073	0.936	12.8	57	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	66	0.00
17	Pyrene	1.326	1.115	15.9	56	0.00
18 S	Terphenyl-d14	0.749	3.330	-344.6#	293#	0.00
19	Benzo(a)anthracene	4.211	3.439	18.3	54	0.00
20	Chrysene	4.750	3.920	17.5	55	0.00
21	Indeno(1,2,3-cd)pyrene	4.146	2.973	28.3#	44#	0.00
22 I	Perylene-d12	1.000	1.000	0.0	61	0.00
23	Benzo(b)fluoranthene	0.697	0.569	18.4	43#	0.00
24	Benzo(k)fluoranthene	1.274	1.166	8.5	56	0.00
25 C	Benzo(a)pyrene	0.770	0.607	21.2#	46#	0.00
26	Dibenzo(a,h)anthracene	0.829	0.660	20.4	43#	0.00
27	Benzo(g,h,i)perylene	4.001	2.997	25.1#	45#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1