

Data Path : Z:\HPCHEM1\BNA_E\Data\BE013015\
 Data File : BE089453.D
 Acq On : 31 Jan 2015 1:11
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 02 00:59:45 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE013015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 31 05:05:18 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	66	0.00
2 I	Naphthalene-d8	5.000	5.000	0.0	68	0.00
3 S	Nitrobenzene-d5	1.000	4.242	-324.2#	304	0.00
4	Naphthalene	1.000	0.851	14.9	60	0.00
5	2-Methylnaphthalene	1.000	0.840	16.0	57	0.00
6 I	Acenaphthene-d10	5.000	5.000	0.0	66	-0.01
7 S	2-Fluorobiphenyl	1.000	4.988	-398.8#	337	0.00
8	Acenaphthylene	1.000	0.821	17.9	55	0.00
9	Acenaphthene	1.000	0.819	18.1	56	0.00
10	Fluorene	1.000	0.820	18.0	55	0.00
11 I	Phenanthrene-d10	5.000	5.000	0.0	66	0.00
12	Pentachlorophenol	1.000	2.245	-124.5#	145	0.00
13	Phenanthrene	1.000	0.813	18.7	55	0.00
14	Anthracene	1.000	0.836	16.4	55	0.00
15	Fluoranthene	1.000	0.872	12.8	57	0.00
16 I	Chrysene-d12	5.000	5.000	0.0	66	0.00
17	Pyrene	1.000	0.841	15.9	56	0.00
18 S	Terphenyl-d14	1.000	4.448	-344.8#	293	0.00
19	Benzo(a)anthracene	1.000	0.817	18.3	54	0.00
20	Chrysene	1.000	0.825	17.5	55	0.00
21	Indeno(1,2,3-cd)pyrene	1.000	0.717	28.3#	44	0.00
22 I	Perylene-d12	5.000	5.000	0.0	61	0.00
23	Benzo(b)fluoranthene	1.000	0.801	19.9	43	0.00
24	Benzo(k)fluoranthene	1.000	0.916	8.4	56	0.00
25 C	Benzo(a)pyrene	1.000	0.775	22.5#	46	0.00
26	Dibenzo(a,h)anthracene	1.000	0.758	24.2	43	0.00
27	Benzo(g,h,i)perylene	1.000	0.749	25.1#	45	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1