

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013018\
 Data File : BE095289.D
 Acq On : 30 Jan 2018 17:14
 Operator : SJ/JU
 Sample : J1244-03DL 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6B44DL

Manual Integrations
 APPROVED

Sohil
 1/31/2018 8:28:14 PM

Quant Time: Jan 31 03:22:02 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 31 01:22:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	2163	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	6668	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	4232	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	9013	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	9356	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	9177	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	244	0.02	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	603	0.02	ng/ul	0.00
Target Compounds						
3) Naphthalene	11.31	128	12846	0.688	ng/ul#	88
5) 2-Methylnaphthalene	12.89	142	10660	0.855	ng/ul	100
8) Acenaphthene	15.08	153	5190	0.329	ng/ul	93
9) Fluorene	16.03	166	3663	0.225	ng/ul#	75
11) Pentachlorophenol	17.36	266	630	0.361	ng/ul	97
12) Phenanthrene	17.75	178	14666	0.484	ng/ul#	88
13) Anthracene	17.85	178	1735	0.063	ng/ul	95
15) Fluoranthene	19.73	202	4943	0.126	ng/ul	84
17) Pyrene	20.08	202	3100m	0.099	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013018\
Data File : BE095289.D
Acq On : 30 Jan 2018 17:14
Operator : SJ/JU
Sample : J1244-03DL 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6B44DL

Manual Integrations
APPROVED

Sohil
1/31/2018 8:28:14 PM

Quant Time: Jan 31 03:22:02 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 01:22:09 2018
Response via : Initial Calibration

