

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013018\
 Data File : BE095309.D
 Acq On : 31 Jan 2018 10:12
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.426

Quant Time: Jan 31 11:47:13 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 31 08:16:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	73	0.00
3	Naphthalene	1.120	1.118	0.2	73	0.00
4 SURR	2-Methylnaphthalene-d10	0.651	0.687	-5.5	76	0.00
5	2-Methylnaphthalene	0.748	0.797	-6.6	78	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
7	Acenaphthylene	1.925	2.076	-7.8	87	0.00
8 C	Acenaphthene	1.493	1.520	-1.8	81	0.00
9	Fluorene	1.540	1.704	-10.6	88	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
11	Pentachlorophenol	0.077	0.098	-27.3#	103	0.00
12	Phenanthrene	1.345	1.357	-0.9	77	0.00
13	Anthracene	1.231	1.310	-6.4	82	0.00
14 SURR	Fluoranthene-d10	1.531	1.414	7.6	73	0.00
15 C	Fluoranthene	1.742	1.573	9.7	72	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	64	0.00
17	Pyrene	1.342	1.691	-26.0#	83	0.00
18	Benzo(a)anthracene	1.251	1.363	-9.0	69	0.00
19	Chrysene	1.299	1.299	0.0	62	0.00
20 I	Perylene-d12	1.000	1.000	0.0	69	0.00
21	Benzo(b)fluoranthene	1.411	1.332	5.6	65	0.00
22	Benzo(k)fluoranthene	1.291	1.277	1.1	68	0.00
23 C	Benzo(a)pyrene	1.260	1.264	-0.3	70	0.00
24	Indeno(1,2,3-cd)pyrene	1.382	1.415	-2.4	71	0.00
25	Dibenzo(a,h)anthracene	1.130	1.158	-2.5	70	0.00
26	Benzo(a,h,i)perylene	1.170	1.184	-1.2	69	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0