

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013018\
 Data File : BE095320.D
 Acq On : 31 Jan 2018 16:46
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.427

Quant Time: Jan 31 17:30:41 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 31 11:55:47 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	76	0.00
3	Naphthalene	1.120	1.122	-0.2	77	0.00
4 SURR	2-Methylnaphthalene-d10	0.651	0.688	-5.7	80	0.00
5	2-Methylnaphthalene	0.748	0.801	-7.1	82	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
7	Acenaphthylene	1.925	2.064	-7.2	90	0.00
8 C	Acenaphthene	1.493	1.482	0.7	83	0.00
9	Fluorene	1.540	1.614	-4.8	88	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	77	0.00
11	Pentachlorophenol	0.077	0.093	-20.8#	97	0.00
12	Phenanthrene	1.345	1.341	0.3	75	0.00
13	Anthracene	1.231	1.314	-6.7	81	0.00
14 SURR	Fluoranthene-d10	1.531	1.382	9.7	71	0.00
15 C	Fluoranthene	1.742	1.535	11.9	69	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	62	0.00
17	Pyrene	1.342	1.592	-18.6	76	0.00
18	Benzo(a)anthracene	1.251	1.372	-9.7	67	0.00
19	Chrysene	1.299	1.290	0.7	60	0.00
20 I	Perylene-d12	1.000	1.000	0.0	68	0.00
21	Benzo(b)fluoranthene	1.411	1.294	8.3	62	0.00
22	Benzo(k)fluoranthene	1.291	1.287	0.3	67	0.00
23 C	Benzo(a)pyrene	1.260	1.249	0.9	67	0.00
24	Indeno(1,2,3-cd)pyrene	1.382	1.410	-2.0	69	0.00
25	Dibenzo(a,h)anthracene	1.130	1.147	-1.5	68	0.00
26	Benzo(a,h,i)perylene	1.170	1.181	-0.9	68	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0