

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013018\
 Data File : BE095310.D
 Acq On : 31 Jan 2018 10:48
 Operator : SJ/JU
 Sample : J1243-01DL 5X
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6B25DL

Quant Time: Jan 31 11:57:04 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 31 08:16:13 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	2739	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	8722	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	5370	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	11762	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	11681	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	12015	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	1278	0.09	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	3259	0.07	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.31	128	36411	1.491	ng/ul#	88
5) 2-Methylnaphthalene	12.89	142	11105	0.681	ng/ul	99
8) Acenaphthene	15.08	153	3066	0.153	ng/ul	94
9) Fluorene	16.02	166	2455	0.119	ng/ul#	81
11) Pentachlorophenol	17.36	266	102212	44.866	ng/ul	96
12) Phenanthrene	17.75	178	10423	0.263	ng/ul#	89
13) Anthracene	17.85	178	1551	0.043	ng/ul#	94
15) Fluoranthene	19.72	202	3889	0.076	ng/ul	85
17) Pyrene	20.08	202	3210	0.082	ng/ul#	83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013018\
Data File : BE095310.D
Acq On : 31 Jan 2018 10:48
Operator : SJ/JU
Sample : J1243-01DL 5X
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6B25DL

Quant Time: Jan 31 11:57:04 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 08:16:13 2018
Response via : Initial Calibration





#3

Naphthalene

Concen: 1.491 ng/ul

RT: 11.31 min Scan# 1348

Delta R.T. -0.00 min

Lab File: BE095310.D

Acq: 31 Jan 2018 10:48

Instrument :

BNA_E

ClientSampleId :

F6B25DL


Tot Ion:128 Resp: 36411

Ion Ratio Lower Upper

128 100

129 11.1 11.3 16.9#




127 13.3 4.0 6.0#



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#5

2-Methylnaphthalene

Concen: 0.681 ng/ul

RT: 12.89 min Scan# 1539

Delta R.T. -0.00 min

Lab File: BE095310.D


Acq: 31 Jan 2018 10:48

Tgt Ion:142 Resp: 11105

Ion Ratio Lower Upper



142 100


141 90.0 72.6 108.8



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E





#8

Acenaphthene

Concen: 0.153 ng/ul

RT: 15.08 min Scan# 1825

Delta R.T. -0.00 min

Lab File: BE095310.D

Acq: 31 Jan 2018 10:48

Instrument :

BNA_E

ClientSampleId :

F6B25DL


Tot Ion:153 Resp: 3066

Ion Ratio Lower Upper

153 100

152 48.6 36.0 54.0




154 83.7 72.0 108.0



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



#9

Fluorene

Concen: 0.119 ng/ul

RT: 16.02 min Scan# 1939

Delta R.T. 0.00 min

Lab File: BE095310.D

Acq: 31 Jan 2018 10:48


Tgt Ion:166 Resp: 2455

Ion Ratio Lower Upper

166 100

165 112.7 73.4 110.2#



167 15.4 14.6 22.0




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#11
 Pentachlorophenol
 Concen: 44.866 ng/ul
 RT: 17.36 min Scan# 1997
 Delta R.T. 0.00 min
 Lab File: BE095310.D
 Acq: 31 Jan 2018 10:48




Instrument :
 BNA_E
 ClientSampleId :
 F6B25DL


Tot Ion:266 Resp: 102212
 Ion Ratio Lower Upper
 266 100
 264 65.5 47.9 71.9
 268 61.2 49.8 74.8



#12
 Phenanthrene
 Concen: 0.263 ng/ul
 RT: 17.75 min Scan# 2014
 Delta R.T. 0.00 min
 Lab File: BE095310.D
 Acq: 31 Jan 2018 10:48

Tgt Ion:178 Resp: 10423
 Ion Ratio Lower Upper
 178 100
 179 16.1 18.4 27.6#
 176 19.8 13.6 20.4





#13

Anthracene

Concen: 0.043 ng/ul

RT: 17.85 min Scan# 2018

Delta R.T. 0.00 min

Lab File: BE095310.D

Acq: 31 Jan 2018 10:48

Instrument :

BNA_E

ClientSampleId :

F6B25DL


Tot Ion:178 Resp: 1551

Ion Ratio Lower Upper

178 100

179 21.1 18.1 27.1




176 22.3 14.7 22.1#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#15

Fluoranthene

Concen: 0.076 ng/ul

RT: 19.72 min Scan# 2395

Delta R.T. 0.00 min

Lab File: BE095310.D

Acq: 31 Jan 2018 10:48


Tgt Ion:202 Resp: 3889

Ion Ratio Lower Upper

202 100

101 11.2 0.0 30.3



100 9.4 0.0 39.5






Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): E





#17

Pvrene

Concen: 0.082 ng/ul

RT: 20.08 min Scan# 2464

Delta R.T. -0.00 min

Lab File: BE095310.D

Acq: 31 Jan 2018 10:48

Instrument :

BNA_E

ClientSampleId :

F6B25DL

Tot Ion:202 Resp: 3210

Ion Ratio Lower Upper

202 100

101 13.8 18.2 27.4#

100 12.8 15.6 23.4#

