

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013018\
 Data File : BE095320.D
 Acq On : 31 Jan 2018 16:46
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.427

Quant Time: Jan 31 17:30:41 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 31 11:55:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	2402	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	7358	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	4384	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	9310	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	9724	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	9427	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	5061	0.42	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	12863	0.36	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.31	128	8257	0.401	ng/ul#	88
5) 2-Methylnaphthalene	12.89	142	5897	0.429	ng/ul	100
7) Acenaphthylene	14.74	152	9049	0.429	ng/ul	96
8) Acenaphthene	15.08	153	6499	0.397	ng/ul	95
9) Fluorene	16.03	166	7077	0.419	ng/ul#	80
11) Pentachlorophenol	17.36	266	868	0.481	ng/ul	96
12) Phenanthrene	17.75	178	12485	0.399	ng/ul#	88
13) Anthracene	17.85	178	12234	0.427	ng/ul#	92
15) Fluoranthene	19.72	202	14292	0.353	ng/ul	84
17) Pyrene	20.08	202	15476	0.474	ng/ul#	79
18) Benzo(a)anthracene	21.79	228	13341	0.439	ng/ul#	86
19) Chrysene	21.84	228	12548	0.397	ng/ul	96
21) Benzo(b)fluoranthene	23.62	252	12200	0.367	ng/ul	87
22) Benzo(k)fluoranthene	23.67	252	12128	0.399	ng/ul#	89
23) Benzo(a)pyrene	24.32	252	11778	0.397	ng/ul#	81
24) Indeno(1,2,3-cd)pyrene	27.25	276	13292	0.408	ng/ul#	81
25) Dibenzo(a,h)anthracene	27.27	278	10816	0.406	ng/ul#	90
26) Benzo(g,h,i)perylene	28.12	276	11129	0.404	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013018\
Data File : BE095320.D
Acq On : 31 Jan 2018 16:46
Operator : SJ/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.427

Quant Time: Jan 31 17:30:41 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 11:55:47 2018
Response via : Initial Calibration

