

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013117\
 Data File : BE092690.D
 Acq On : 31 Jan 2017 17:06
 Operator : SJ/MA
 Sample : PB96577BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB96577BL

Quant Time: Feb 01 02:49:18 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	10717	5.00	ng	0.00
7) Naphthalene-d8	10.91	136	47483	5.00	ng	0.00
12) Acenaphthene-d10	14.78	164	29072	5.00	ng	0.00
18) Phenanthrene-d10	17.53	188	59723	5.00	ng	0.00
24) Chrysene-d12	21.72	240	70227	5.00	ng	0.00
31) Perylene-d12	24.06	264	77387	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.64	112	24642	10.30	ng	-0.01
5) Phenol-d6	7.31	99	29072	9.91	ng	-0.02
8) Nitrobenzene-d5	9.32	82	12866	4.17	ng	-0.02
13) 2,4,6-Tribromophenol	16.31	330	4344	5.11	ng	0.01
14) 2-Fluorobiphenyl	13.41	172	31534	4.41	ng	-0.01
26) Terphenyl-d14	20.14	244	44383	4.47	ng	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013117\
Data File : BE092690.D
Acq On : 31 Jan 2017 17:06
Operator : SJ/MA
Sample : PB96577BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB96577BL

Quant Time: Feb 01 02:49:18 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 17 10:14:01 2017
Response via : Initial Calibration

