

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013117\
 Data File : BE092705.D
 Acq On : 1 Feb 2017 2:27
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 01 03:24:19 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	195	0.00
2	1,4-Dioxane	0.400	0.330	17.5	159	0.01
3	n-Nitrosodimethylamine	0.400	0.277	30.8#	154	0.01
4 S	2-Fluorophenol	0.400	0.405	-1.3	232	0.00
5 S	Phenol-d6	0.400	0.463	-15.8	274	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.342	14.5	199	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	206	0.00
8 S	Nitrobenzene-d5	0.400	0.371	7.3	231	-0.02
9	Naphthalene	0.400	0.394	1.5	220	-0.02
10	Hexachlorobutadiene	0.400	0.416	-4.0	226	-0.02
11	2-Methylnaphthalene	0.400	0.402	-0.5	232	-0.02
12 I	Acenaphthene-d10	5.000	5.000	0.0	214	0.00
13 S	2,4,6-Tribromophenol	0.400	0.483	-20.7	282	-0.01
14 S	2-Fluorobiphenyl	0.400	0.386	3.5	223	-0.02
15	Acenaphthylene	0.400	0.409	-2.2	247	-0.02
16	Acenaphthene	0.400	0.402	-0.5	227	0.00
17	Fluorene	0.400	0.389	2.8	232	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	208	0.00
19	Hexachlorobenzene	0.400	0.363	9.3	193	0.00
20	Pentachlorophenol	0.400	0.408	-2.0	232	0.00
21	Phenanthrene	0.400	0.398	0.5	214	0.00
22	Anthracene	0.400	0.417	-4.2	243	-0.02
23	Fluoranthene	0.400	0.407	-1.7	233	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	171	0.00
25	Pyrene	0.400	0.457	-14.2	234	-0.02
26 S	Terphenyl-d14	0.400	0.440	-10.0	219	0.00
27	Benzo(a)anthracene	0.400	0.430	-7.5	213	0.00
28	Chrysene	0.400	0.421	-5.2	193	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.744	-86.0#	356	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.400	0.0	195	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	167	0.00
32	Benzo(b)fluoranthene	0.400	0.420	-5.0	191	-0.01
33	Benzo(k)fluoranthene	0.400	0.425	-6.2	207	-0.01
34 C	Benzo(a)pyrene	0.400	0.425	-6.2	206	-0.01
35	Dibenzo(a,h)anthracene	0.400	0.417	-4.2	197	-0.03
36	Benzo(g,h,i)perylene	0.400	0.407	-1.7	188	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0