

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013117\
 Data File : BE092689.D
 Acq On : 31 Jan 2017 16:27
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad
 2/1/2017 2:28:29 PM

Quant Time: Feb 01 10:34:49 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	15174	5.00	ng	0.00
7) Naphthalene-d8	10.91	136	72894	5.00	ng	0.00
12) Acenaphthene-d10	14.78	164	45970	5.00	ng	0.00
18) Phenanthrene-d10	17.53	188	104240	5.00	ng	0.00
24) Chrysene-d12	21.72	240	107921	5.00	ng	0.00
31) Perylene-d12	24.06	264	125048	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.65	112	1219	0.36	ng	0.00
5) Phenol-d6	7.31	99	1625	0.39	ng	-0.02
8) Nitrobenzene-d5	9.32	82	1649	0.35	ng	-0.02
13) 2,4,6-Tribromophenol	16.32	330	452	0.46	ng	0.02
14) 2-Fluorobiphenyl	13.41	172	4478	0.40	ng	-0.01
26) Terphenyl-d14	20.16	244	6225	0.41	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.37	88	855	0.35	ng	98
3) n-Nitrosodimethylamine	3.77	42	688	0.25	ng	# 99
6) bis(2-Chloroethyl)ether	7.57	93	1564	0.35	ng	# 80
9) Naphthalene	10.95	128	5904	0.39	ng	96
10) Hexachlorobutadiene	11.24	225	935	0.42	ng	99
11) 2-Methylnaphthalene	12.60	142	3597	0.38	ng	92
15) Acenaphthylene	14.49	152	25419	0.40	ng	100
16) Acenaphthene	14.85	154	4091	0.40	ng	97
17) Fluorene	15.83	166	4945	0.39	ng	99
19) Hexachlorobenzene	16.87	284	1415m	0.33	ng	
20) Pentachlorophenol	17.51	266	318m	0.30	ng	
21) Phenanthrene	17.58	178	8634	0.38	ng	95
22) Anthracene	17.67	178	8260	0.39	ng	99
23) Fluoranthene	19.58	202	41641	0.41	ng	98
25) Pyrene	19.93	202	44825	0.42	ng	99
27) Benzo(a)anthracene	21.70	228	45212	0.43	ng	# 76
28) Chrysene	21.75	228	52120m	0.44	ng	
29) Bis(2-ethylhexyl)phthalate	21.61	149	9557	0.64	ng	# 71
30) Indeno(1,2,3-cd)pyrene	26.53	276	56751	0.44	ng	97
32) Benzo(b)fluoranthene	23.34	252	12099	0.40	ng	100
33) Benzo(k)fluoranthene	23.38	252	12781	0.42	ng	96
34) Benzo(a)pyrene	23.95	252	11853	0.42	ng	97
35) Dibenzo(a,h)anthracene	26.55	278	11516	0.41	ng	# 94
36) Benzo(g,h,i)perylene	27.30	276	48427	0.41	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE013117\
Data File : BE092689.D
Acq On : 31 Jan 2017 16:27
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM

Quant Time: Feb 01 10:34:49 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 17 10:14:01 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.14 min Scan# 575

Delta R.T. 0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
152	100		
150	132.8	106.0	159.0
115	39.0	31.2	46.8

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



#2

1,4-Dioxane

Concen: 0.35 ng




RT: 3.37 min Scan# 68


Delta R.T. 0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion	Ratio	Lower	Upper
88	100		
58	80.8	63.6	95.4
43	37.0	28.0	42.0





#3





n-Nitrosodimethylamine
Concen: 0.25 ng
RT: 3.77 min Scan# 102
Delta R.T. 0.02 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
42	100		
74	108.4	86.0	129.0
44	3.8	5.1	7.7#

Manual Integrations
APPROVED




mohammad
2/1/2017 2:28:29 PM




#4

2-Fluorophenol
Concen: 0.36 ng
RT: 5.65 min Scan# 360
Delta R.T. -0.00 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Tgt Ion	Ratio	Lower	Upper
112	100		
64	65.7	53.5	80.3
63	34.7	27.9	41.9





#5

Phenol-d6

Concen: 0.39 ng

RT: 7.31 min Scan# 539

Delta R.T. -0.02 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	99	Resp:	1625
Ion	Ratio	Lower	Upper
99	100		
42	27.1	16.0	24.0#
71	36.9	30.6	45.8

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.35 ng




RT: 7.57 min Scan# 550


Delta R.T. 0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion:	93	Resp:	1564
Ion	Ratio	Lower	Upper
93	100		
63	65.3	71.6	107.4#
95	28.5	24.0	36.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.91 min Scan# 706

Delta R.T. -0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	136	Resp:	72894
Ion Ratio	Lower	Upper	
136	100		
137	11.3	8.2	12.4
54	6.7	9.6	14.4#
68	5.0	6.7	10.1#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



#8

Nitrobenzene-d5

Concen: 0.35 ng




RT: 9.32 min Scan# 626


Delta R.T. -0.02 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion:	82	Resp:	1649
Ion Ratio	Lower	Upper	
82	100		
128	35.2	39.0	58.4#
54	63.4	44.8	67.2





#9

Naphthalene

Concen: 0.39 ng

RT: 10.95 min Scan# 708

Delta R.T. -0.02 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	128	Resp:	5904
Ion	Ratio	Lower	Upper
128	100		
129	12.0	11.2	16.8
127	15.9	11.6	17.4

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



#10

Hexachlorobutadiene

Concen: 0.42 ng




RT: 11.24 min Scan# 723


Delta R.T. -0.02 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion:	225	Resp:	935
Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.6	76.0
227	63.9	51.4	77.0






#11
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.60 min Scan# 812
Delta R.T. -0.01 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4




Tot Ion:142 Resp: 3597
Ion Ratio Lower Upper
142 100
141 86.4 73.7 110.5
115 35.4 34.0 51.0

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM




Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E






#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.78 min Scan# 978
Delta R.T. -0.00 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Tgt Ion:164 Resp: 45970
Ion Ratio Lower Upper
164 100
162 107.4 71.4 107.0#
160 49.1 27.4 41.2#



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E









#13
 2,4,6-Tribromophenol
 Concen: 0.46 ng
 RT: 16.32 min Scan# 1089
 Delta R.T. 0.02 min
 Lab File: BE092689.D
 Acq: 31 Jan 2017 16:27

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Tot Ion:330 Resp: 452
 Ion Ratio Lower Upper
 330 100
 332 96.2 75.4 113.0
 141 40.5 31.0 46.6




Manual Integrations
 APPROVED


mohammad
 2/1/2017 2:28:29 PM



#14
 2-Fluorobiphenyl
 Concen: 0.40 ng
 RT: 13.41 min Scan# 882
 Delta R.T. -0.01 min
 Lab File: BE092689.D
 Acq: 31 Jan 2017 16:27

Tgt Ion:172 Resp: 4478
 Ion Ratio Lower Upper
 172 100
 171 34.3 30.1 45.1
 170 23.1 21.8 32.6





#15

Acenaphthylene

Concen: 0.40 ng

RT: 14.49 min Scan# 961

Delta R.T. -0.02 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	152	Resp:	25419
Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	13.0	10.4	15.6

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



#16

Acenaphthene

Concen: 0.40 ng




RT: 14.85 min Scan# 982


Delta R.T. -0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion:	154	Resp:	4091
Ion	Ratio	Lower	Upper
154	100		
153	438.7	350.8	526.2
152	226.3	171.1	256.7









#17
Fluorene
 Concen: 0.39 ng
 RT: 15.83 min Scan# 1047
 Delta R.T. -0.01 min
 Lab File: BE092689.D
 Acq: 31 Jan 2017 16:27

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
166	100		
165	99.3	78.2	117.4
167	13.3	11.0	16.4




Manual Integrations
 APPROVED


mohammad
 2/1/2017 2:28:29 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.53 min Scan# 1152
 Delta R.T. 0.00 min
 Lab File: BE092689.D
 Acq: 31 Jan 2017 16:27

Tgt Ion	Ratio	Lower	Upper
188	100		
94	7.0	3.2	4.8#
80	6.3	2.6	3.8#









#19
Hexachlorobenzene
Concen: 0.33 ng m
RT: 16.87 min Scan# 1123
Delta R.T. 0.00 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Resp	Lower	Upper
284	100			
142	0.0	29.1	43.7#	
249	0.0	27.9	41.9#	




Manual Integrations
APPROVED


mohammad
2/1/2017 2:28:29 PM



#20
Pentachlorophenol
Concen: 0.30 ng m
RT: 17.51 min Scan# 1151
Delta R.T. 0.28 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Tgt Ion	Ratio	Resp	Lower	Upper
266	100			
268	88.9	62.5	93.7	
264	87.3	57.2	85.8#	









#21
Phenanthrene
Concen: 0.38 ng
RT: 17.58 min Scan# 1154
Delta R.T. 0.00 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.8	23.8
179	16.7	16.0	24.0




Manual Integrations
APPROVED


mohammad
2/1/2017 2:28:29 PM



#22
Anthracene
Concen: 0.39 ng
RT: 17.67 min Scan# 1158
Delta R.T. -0.02 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.3	15.0	22.4
179	15.9	12.2	18.2





#23

Fluoranthene

Concen: 0.41 ng

RT: 19.58 min Scan# 1264

Delta R.T. -0.02 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	202	Resp:	41641
Ion Ratio	Lower	Upper	
202	100		
101	2.8	2.2	3.4
203	18.0	13.7	20.5

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.72 min Scan# 1374


Delta R.T. 0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion:	240	Resp:	107921
Ion Ratio	Lower	Upper	
240	100		
120	0.9	2.6	4.0#
236	21.4	24.6	37.0#





#25

Pvrene

Concen: 0.42 ng

RT: 19.93 min Scan# 1285

Delta R.T. -0.02 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	202	Resp:	44825
Ion Ratio	Lower	Upper	
202	100		
200	5.1	4.2	6.4
203	17.6	13.9	20.9

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



#26

Terphenyl-d14

Concen: 0.41 ng




RT: 20.16 min Scan# 1298


Delta R.T. 0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion:	244	Resp:	6225
Ion Ratio	Lower	Upper	
244	100		
212	8.6	6.7	10.1
122	5.0	3.6	5.4





#27

Benzo(a)anthracene

Concen: 0.43 ng

RT: 21.70 min Scan# 1373

Delta R.T. 0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	228	Resp:	45212
Ion	Ratio	Lower	Upper
228	100		
226	17.9	23.6	35.4#
229	28.5	13.3	19.9#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



#28

Chrysene

Concen: 0.44 ng m




RT: 21.75 min Scan# 1375


Delta R.T. 0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion:	228	Resp:	52120
Ion	Ratio	Lower	Upper
228	100		
226	24.5	36.3	54.5#
229	21.1	10.6	16.0#





#29

Bis(2-ethylhexyl)phthalate

Concen: 0.64 ng

RT: 21.61 min Scan# 1369

Delta R.T. 0.00 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4




Tot Ion:	149	Resp:	9557
Ion Ratio	Lower	Upper	
149	100		
167	26.5	16.0	24.0#
279	0.0	16.0	24.0#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E



#30

Indeno(1,2,3-cd)pyrene

Concen: 0.44 ng

RT: 26.53 min Scan# 1712




Delta R.T. -0.02 min


Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion:	276	Resp:	56751
Ion Ratio	Lower	Upper	
276	100		
138	28.1	20.3	30.5
277	24.8	19.7	29.5

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.06 min Scan# 1542
Delta R.T. -0.00 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion	Ratio	Resp	Lower	Upper
264	100	125048		
260	25.1		20.1	30.1
265	22.2		17.9	26.9




Manual Integrations
APPROVED


mohammad
2/1/2017 2:28:29 PM



#32
Benzo(b)fluoranthene
Concen: 0.40 ng
RT: 23.34 min Scan# 1479
Delta R.T. -0.01 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	12099		
253	23.3		18.7	28.1
125	13.1		10.5	15.7









#33
Benzo(k)fluoranthene
Concen: 0.42 ng
RT: 23.38 min Scan# 1483
Delta R.T. -0.01 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tot Ion:	252	Resp:	12781
Ion Ratio	Lower	Upper	
252	100		
253	23.1	20.3	30.5
125	12.9	9.6	14.4




Manual Integrations
APPROVED


mohammad
2/1/2017 2:28:29 PM



#34
Benzo(a)pyrene
Concen: 0.42 ng
RT: 23.95 min Scan# 1532
Delta R.T. -0.01 min
Lab File: BE092689.D
Acq: 31 Jan 2017 16:27

Tgt Ion:	252	Resp:	11853
Ion Ratio	Lower	Upper	
252	100		
253	23.2	19.0	28.6
125	17.6	11.8	17.8





#35

Dibenzo(a,h)anthracene

Concen: 0.41 ng

RT: 26.55 min Scan# 1713

Delta R.T. -0.03 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4

Tot Ion:	278	Resp:	11516
Ion Ratio	Lower	Upper	
278	100		
139	21.0	13.8	20.8#
279	24.7	21.4	32.2

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:29 PM



#36

Benzo(g,h,i)perylene

Concen: 0.41 ng

RT: 27.30 min Scan# 1757

Delta R.T. -0.02 min

Lab File: BE092689.D

Acq: 31 Jan 2017 16:27

Tgt Ion:	276	Resp:	48427
Ion Ratio	Lower	Upper	
276	100		
277	25.3	19.8	29.8
138	23.1	19.8	29.6

