

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013117\
 Data File : BE092693.D
 Acq On : 31 Jan 2017 19:00
 Operator : SJ/MA
 Sample : PB96580BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB96580BS

Manual Integrations
 APPROVED

mohammad
 2/1/2017 2:28:35 PM


Quant Time: Feb 01 10:57:23 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	10562	5.00	ng	0.00
7) Naphthalene-d8	10.91	136	47492	5.00	ng	0.00
12) Acenaphthene-d10	14.78	164	29111	5.00	ng	0.00
18) Phenanthrene-d10	17.53	188	61792	5.00	ng	0.00
24) Chrysene-d12	21.72	240	65316	5.00	ng	0.00
31) Perylene-d12	24.07	264	74524	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.64	112	20996	8.90	ng	-0.02
5) Phenol-d6	7.31	99	26588	9.20	ng	-0.02
8) Nitrobenzene-d5	9.29	82	13362	4.33	ng	-0.05
13) 2,4,6-Tribromophenol	16.28	330	7028	8.17	ng	-0.01
14) 2-Fluorobiphenyl	13.40	172	28598	3.99	ng	-0.02
26) Terphenyl-d14	20.14	244	38613	4.18	ng	-0.02

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.38	88	4714	3.30	ng	# 73
3) n-Nitrosodimethylamine	3.72	42	7405	3.89	ng	# 97
6) bis(2-Chloroethyl)ether	7.54	93	13192	4.30	ng	93
9) Naphthalene	10.95	128	40273	4.06	ng	# 95
10) Hexachlorobutadiene	11.24	225	6142	4.19	ng	99
11) 2-Methylnaphthalene	12.59	142	25737	4.18	ng	# 87
15) Acenaphthylene	14.49	152	168025	4.19	ng	100
16) Acenaphthene	14.85	154	25389	3.95	ng	94
17) Fluorene	15.83	166	33142	4.10	ng	100
19) Hexachlorobenzene	16.87	284	8502	4.20	ng	# 39
20) Pentachlorophenol	17.30	266	4865m	7.66	ng	
21) Phenanthrene	17.58	178	54552	4.00	ng	# 94
22) Anthracene	17.67	178	55775	4.43	ng	99
23) Fluoranthene	19.58	202	260145	4.30	ng	99
25) Pyrene	19.93	202	281725	4.38	ng	99
27) Benzo(a)anthracene	21.70	228	275118	4.28	ng	# 73
28) Chrysene	21.75	228	303712m	4.20	ng	
29) Bis(2-ethylhexyl)phthalate	21.61	149	54509	4.33	ng	# 72
30) Indeno(1,2,3-cd)pyrene	26.53	276	366270	4.68	ng	98
32) Benzo(b)fluoranthene	23.34	252	80299	4.45	ng	# 96
33) Benzo(k)fluoranthene	23.38	252	77709	4.29	ng	94
34) Benzo(a)pyrene	23.95	252	78615	4.65	ng	# 94
35) Dibenzo(a,h)anthracene	26.55	278	75163	4.54	ng	95
36) Benzo(g,h,i)perylene	27.30	276	313410	4.48	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.14 min Scan# 575

Delta R.T. 0.00 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Instrument :

BNA_E





ClientSampleId :

PB96580BS

Tot Ion	Ratio	Lower	Upper
152	100		
150	134.8	106.0	159.0
115	38.1	31.2	46.8

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#2

1,4-Dioxane

Concen: 3.30 ng




RT: 3.38 min Scan# 68


Delta R.T. 0.00 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
88	100		
58	92.6	63.6	95.4
43	66.5	28.0	42.0





#3

n-Nitrosodimethylamine

Concen: 3.89 ng

RT: 3.72 min Scan# 98

Delta R.T. -0.02 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Instrument :

BNA_E





ClientSampleId :

PB96580BS

Tot Ion:	42	Resp:	7405
Ion Ratio	Lower	Upper	
42	100		
74	105.0	86.0	129.0
44	10.2	5.1	7.7#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#4

2-Fluorophenol

Concen: 8.90 ng




RT: 5.64 min Scan# 357


Delta R.T. -0.02 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Tgt Ion:	112	Resp:	20996
Ion Ratio	Lower	Upper	
112	100		
64	68.8	53.5	80.3
63	36.7	27.9	41.9





#5

Phenol-d6

Concen: 9.20 ng

RT: 7.31 min Scan# 539

Delta R.T. -0.02 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Instrument :

BNA_E





ClientSampleId :

PB96580BS

Tot Ion:	99	Resp:	26588
Ion Ratio	Lower	Upper	
99	100		
42	0.0	16.0	24.0#
71	0.0	30.6	45.8#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#6

bis(2-Chloroethyl)ether

Concen: 4.30 ng




RT: 7.54 min Scan# 549

Delta R.T. -0.02 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Tgt Ion:	93	Resp:	13192
Ion Ratio	Lower	Upper	
93	100		
63	81.2	71.6	107.4
95	30.6	24.0	36.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.91 min Scan# 706

Delta R.T. 0.00 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Instrument :

BNA_E





ClientSampleId :

PB96580BS

Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	11.5	8.2	12.4	
54	6.1	9.6	14.4	
68	4.7	6.7	10.1	

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#8

Nitrobenzene-d5

Concen: 4.33 ng




RT: 9.29 min Scan# 625

Delta R.T. -0.05 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	18.9	39.0	58.4	
54	65.5	44.8	67.2	





#9

Naphthalene

Concen: 4.06 ng

RT: 10.95 min Scan# 708

Delta R.T. -0.02 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Instrument :

BNA_E





ClientSampleId :

PB96580BS

Tot Ion:	128	Resp:	40273
Ion Ratio	Lower	Upper	
128	100		
129	10.6	11.2	16.8#
127	13.8	11.6	17.4

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#10

Hexachlorobutadiene

Concen: 4.19 ng




RT: 11.24 min Scan# 723


Delta R.T. -0.02 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Tgt Ion:	225	Resp:	6142
Ion Ratio	Lower	Upper	
225	100		
223	62.8	50.6	76.0
227	63.4	51.4	77.0






#11
2-Methylnaphthalene
Concen: 4.18 ng
RT: 12.59 min Scan# 811
Delta R.T. -0.02 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Instrument :
BNA_E
ClientSampleId :
PB96580BS

Tot Ion	Ratio	Lower	Upper
142	100		
141	81.9	73.7	110.5
115	30.9	34.0	51.0#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM






Abundance

Ion 142.00 (141.70 to 142.70): E


Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.78 min Scan# 978
Delta R.T. 0.00 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.2	71.4	107.0
160	44.7	27.4	41.2#






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#13
 2,4,6-Tribromophenol
 Concen: 8.17 ng
 RT: 16.28 min Scan# 1086
 Delta R.T. -0.01 min
 Lab File: BE092693.D
 Acq: 31 Jan 2017 19:00

Instrument :
 BNA_E
 ClientSampleId :
 PB96580BS

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.5	75.4	113.0
141	43.7	31.0	46.6




Manual Integrations
 APPROVED

mohammad
 2/1/2017 2:28:35 PM



#14
 2-Fluorobiphenyl
 Concen: 3.99 ng
 RT: 13.40 min Scan# 881
 Delta R.T. -0.02 min
 Lab File: BE092693.D
 Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.8	30.1	45.1
170	23.4	21.8	32.6









#15
Acenaphthylene
Concen: 4.19 ng
RT: 14.49 min Scan# 961
Delta R.T. -0.02 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Instrument :
BNA_E
ClientSampleId :
PB96580BS

Tot Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	13.1	10.4	15.6




Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#16
Acenaphthene
Concen: 3.95 ng
RT: 14.85 min Scan# 982
Delta R.T. 0.00 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
154	100		
153	453.3	350.8	526.2
152	222.9	171.1	256.7









#17
Fluorene
Concen: 4.10 ng
RT: 15.83 min Scan# 1047
Delta R.T. -0.01 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Instrument :
BNA_E
ClientSampleId :
PB96580BS

Tot Ion	Ratio	Lower	Upper
166	100		
165	97.5	78.2	117.4
167	13.4	11.0	16.4




Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#18
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.53 min Scan# 1152
Delta R.T. 0.00 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
188	100		
94	7.8	3.2	4.8#
80	7.1	2.6	3.8#






#19
Hexachlorobenzene
Concen: 4.20 ng
RT: 16.87 min Scan# 1123
Delta R.T. 0.00 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Instrument :
BNA_E
ClientSampleId :
PB96580BS




Tot Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	0.0	27.9	41.9#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#20
Pentachlorophenol
Concen: 7.66 ng m
RT: 17.30 min Scan# 1142
Delta R.T. 0.07 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
266	100		
268	68.0	62.5	93.7
264	68.0	57.2	85.8



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E





#21

Phenanthrene

Concen: 4.00 ng

RT: 17.58 min Scan# 1154

Delta R.T. 0.00 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Instrument :

BNA_E





ClientSampleId :

PB96580BS

Tot Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.8	23.8
179	15.8	16.0	24.0

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#22

Anthracene

Concen: 4.43 ng




RT: 17.67 min Scan# 1158


Delta R.T. -0.02 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.1	15.0	22.4
179	15.7	12.2	18.2





#23

Fluoranthene

Concen: 4.30 ng

RT: 19.58 min Scan# 1264

Delta R.T. -0.02 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Instrument :

BNA_E





ClientSampleId :

PB96580BS

Tot Ion:	202	Resp:	260145
Ion Ratio	Lower	Upper	
202	100		
101	2.9	2.2	3.4
203	17.8	13.7	20.5

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.72 min Scan# 1374


Delta R.T. 0.00 min

Lab File: BE092693.D

Acq: 31 Jan 2017 19:00

Tgt Ion:	240	Resp:	65316
Ion Ratio	Lower	Upper	
240	100		
120	0.8	2.6	4.0#
236	21.1	24.6	37.0#









#25
Pvrene
Concen: 4.38 ng
RT: 19.93 min Scan# 1285
Delta R.T. -0.02 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Instrument :
BNA_E
ClientSampleId :
PB96580BS

Tot Ion	Ratio	Resp Lower	Upper
202	100		
200	5.3	4.2	6.4
203	17.9	13.9	20.9




Manual Integrations
APPROVED


mohammad
2/1/2017 2:28:35 PM



#26
Terphenyl-d14
Concen: 4.18 ng
RT: 20.14 min Scan# 1297
Delta R.T. -0.02 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Resp Lower	Upper
244	100		
212	9.7	6.7	10.1
122	12.8	3.6	5.4#









#27
Benzo(a)anthracene
Concen: 4.28 ng
RT: 21.70 min Scan# 1373
Delta R.T. 0.00 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Instrument :
BNA_E
ClientSampled :
PB96580BS

Tot Ion	Ratio	Resp	Lower	Upper
228	100	275118		
226	15.8	23.6	35.4#	
229	29.3	13.3	19.9#	




Manual Integrations
APPROVED


mohammad
2/1/2017 2:28:35 PM



#28
Chrysene
Concen: 4.20 ng m
RT: 21.75 min Scan# 1375
Delta R.T. 0.00 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Resp	Lower	Upper
228	100	303712		
226	25.1	36.3	54.5#	
229	20.5	10.6	16.0#	









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 4.33 ng
 RT: 21.61 min Scan# 1369
 Delta R.T. 0.00 min
 Lab File: BE092693.D
 Acq: 31 Jan 2017 19:00

Instrument :
 BNA_E
 ClientSampleId :
 PB96580BS

Tot Ion	Ratio	Lower	Upper
149	100		
167	25.6	16.0	24.0#
279	0.0	16.0	24.0#




Manual Integrations
 APPROVED


mohammad
 2/1/2017 2:28:35 PM



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 4.68 ng
 RT: 26.53 min Scan# 1712
 Delta R.T. -0.02 min
 Lab File: BE092693.D
 Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
276	100		
138	27.4	20.3	30.5
277	24.8	19.7	29.5









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.07 min Scan# 1542
Delta R.T. 0.00 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Instrument :
BNA_E
ClientSampleId :
PB96580BS

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.6	20.1	30.1
265	22.4	17.9	26.9




Manual Integrations
APPROVED


mohammad
2/1/2017 2:28:35 PM



#32
Benzo(b)fluoranthene
Concen: 4.45 ng
RT: 23.34 min Scan# 1479
Delta R.T. -0.01 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.8	18.7	28.1
125	10.1	10.5	15.7









#33
Benzo(k)fluoranthene
Concen: 4.29 ng
RT: 23.38 min Scan# 1483
Delta R.T. -0.01 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Instrument :
BNA_E
ClientSampleId :
PB96580BS

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.0	20.3	30.5
125	10.2	9.6	14.4




Manual Integrations
APPROVED


mohammad
2/1/2017 2:28:35 PM



#34
Benzo(a)pyrene
Concen: 4.65 ng
RT: 23.95 min Scan# 1532
Delta R.T. -0.01 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.4	19.0	28.6
125	11.6	11.8	17.8









#35
Dibenzo(a,h)anthracene
Concen: 4.54 ng
RT: 26.55 min Scan# 1713
Delta R.T. -0.03 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Instrument :
BNA_E
ClientSampleId :
PB96580BS

Tot Ion	Ratio	Lower	Upper
278	100		
139	18.3	13.8	20.8
279	23.1	21.4	32.2

Manual Integrations
APPROVED

mohammad
2/1/2017 2:28:35 PM



#36
Benzo(g,h,i)perylene
Concen: 4.48 ng
RT: 27.30 min Scan# 1757
Delta R.T. -0.02 min
Lab File: BE092693.D
Acq: 31 Jan 2017 19:00

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.2	19.8	29.8
138	21.3	19.8	29.6

