

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013117\
 Data File : BE092705.D
 Acq On : 1 Feb 2017 2:27
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

mohammad
 2/1/2017 2:29:24 PM

Quant Time: Feb 01 03:24:19 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	18995	5.00	ng	0.00
7) Naphthalene-d8	10.91	136	92895	5.00	ng	0.00
12) Acenaphthene-d10	14.78	164	61836	5.00	ng	0.00
18) Phenanthrene-d10	17.53	188	130414	5.00	ng	0.00
24) Chrysene-d12	21.72	240	122162	5.00	ng	0.00
31) Perylene-d12	24.07	264	129450	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.65	112	1718	0.41	ng	0.00
5) Phenol-d6	7.31	99	2405	0.46	ng	-0.02
8) Nitrobenzene-d5	9.32	82	2235	0.37	ng	-0.02
13) 2,4,6-Tribromophenol	16.28	330	654	0.48	ng	-0.01
14) 2-Fluorobiphenyl	13.40	172	5877	0.39	ng	-0.02
26) Terphenyl-d14	20.16	244	7592	0.44	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.39	88	1010	0.33	ng	97
3) n-Nitrosodimethylamine	3.76	42	950	0.28	ng	98
6) bis(2-Chloroethyl)ether	7.57	93	1888	0.34	ng	86
9) Naphthalene	10.95	128	7635	0.39	ng	96
10) Hexachlorobutadiene	11.24	225	1193	0.42	ng	99
11) 2-Methylnaphthalene	12.59	142	4835	0.40	ng	99
15) Acenaphthylene	14.49	152	34881	0.41	ng	100
16) Acenaphthene	14.85	154	5489	0.40	ng	99
17) Fluorene	15.83	166	6666	0.39	ng	99
19) Hexachlorobenzene	16.87	284	1907m	0.36	ng	
20) Pentachlorophenol	17.23	266	547	0.41	ng	97
21) Phenanthrene	17.58	178	11461	0.40	ng	95
22) Anthracene	17.67	178	11075	0.42	ng	99
23) Fluoranthene	19.58	202	51911	0.41	ng	98
25) Pyrene	19.93	202	55076	0.46	ng	99
27) Benzo(a)anthracene	21.70	228	51700	0.43	ng	# 77
28) Chrysene	21.75	228	57035m	0.42	ng	
29) Bis(2-ethylhexyl)phthalate	21.61	149	13294	0.74	ng	# 73
30) Indeno(1,2,3-cd)pyrene	26.53	276	58505	0.40	ng	97
32) Benzo(b)fluoranthene	23.34	252	13168	0.42	ng	98
33) Benzo(k)fluoranthene	23.38	252	13357	0.42	ng	97
34) Benzo(a)pyrene	23.95	252	12507	0.43	ng	98
35) Dibenzo(a,h)anthracene	26.55	278	11992	0.42	ng	# 95
36) Benzo(g,h,i)perylene	27.30	276	49459	0.41	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE013117\
Data File : BE092705.D
Acq On : 1 Feb 2017 2:27
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM

Quant Time: Feb 01 03:24:19 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE011617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 17 10:14:01 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.14 min Scan# 575

Delta R.T. 0.00 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tot Ion:	152	Resp:	18995
Ion	Ratio	Lower	Upper
152	100		
150	133.5	106.0	159.0
115	38.3	31.2	46.8

Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.33 ng


RT: 3.39 min Scan# 69

Delta R.T. 0.01 min



Lab File: BE092705.D


Acq: 1 Feb 2017 2:27

Tgt Ion:	88	Resp:	1010
Ion	Ratio	Lower	Upper
88	100		
58	75.8	63.6	95.4
43	35.0	28.0	42.0



Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0





#3





n-Nitrosodimethylamine
Concen: 0.28 ng
RT: 3.76 min Scan# 101
Delta R.T. 0.01 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
42	100		
74	109.7	86.0	129.0
44	7.2	5.1	7.7

Manual Integrations
APPROVED




mohammad
2/1/2017 2:29:24 PM




#4

2-Fluorophenol
Concen: 0.41 ng
RT: 5.65 min Scan# 360
Delta R.T. -0.00 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Tgt Ion	Ratio	Lower	Upper
112	100		
64	67.1	53.5	80.3
63	36.3	27.9	41.9





#5

Phenol-d6

Concen: 0.46 ng

RT: 7.31 min Scan# 539

Delta R.T. -0.02 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tot Ion:	99	Resp:	2405
Ion Ratio	Lower	Upper	
99	100		
42	27.0	16.0	24.0#
71	35.1	30.6	45.8

Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 0.34 ng


RT: 7.57 min Scan# 550

Delta R.T. 0.00 min



Lab File: BE092705.D


Acq: 1 Feb 2017 2:27

Tgt Ion:	93	Resp:	1888
Ion Ratio	Lower	Upper	
93	100		
63	71.8	71.6	107.4
95	29.6	24.0	36.0



Abundance Ion 93.00 (92.70 to 93.70): BE0
Ion 63.00 (62.70 to 63.70): BE0
Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.91 min Scan# 706

Delta R.T. -0.00 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC

Tot Ion	Ratio	Resp Lower	Upper
136	100		
137	11.3	8.2	12.4
54	6.8	9.6	14.4#
68	5.2	6.7	10.1#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.37 ng

RT: 9.32 min Scan# 626

Delta R.T. -0.02 min

Lab File: BE092705.D




Acq: 1 Feb 2017 2:27

Tgt Ion	Ratio	Resp Lower	Upper
82	100		
128	40.7	39.0	58.4
54	55.2	44.8	67.2

Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 0.39 ng

RT: 10.95 min Scan# 708

Delta R.T. -0.02 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion:	128	Resp:	7635
Ion	Ratio	Lower	Upper
128	100		
129	11.7	11.2	16.8
127	15.3	11.6	17.4

Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM



#10

Hexachlorobutadiene

Concen: 0.42 ng




RT: 11.24 min Scan# 723

Delta R.T. -0.02 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Tgt Ion:	225	Resp:	1193
Ion	Ratio	Lower	Upper
225	100		
223	63.6	50.6	76.0
227	63.0	51.4	77.0









#11
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.59 min Scan# 811
Delta R.T. -0.02 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion:142 Resp: 4835
Ion Ratio Lower Upper
142 100
141 92.5 73.7 110.5
115 41.6 34.0 51.0




Manual Integrations
APPROVED


mohammad
2/1/2017 2:29:24 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.78 min Scan# 978
Delta R.T. 0.00 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Tgt Ion:164 Resp: 61836
Ion Ratio Lower Upper
164 100
162 101.3 71.4 107.0
160 43.6 27.4 41.2#









#13
 2,4,6-Tribromophenol
 Concen: 0.48 ng
 RT: 16.28 min Scan# 1086
 Delta R.T. -0.01 min
 Lab File: BE092705.D
 Acq: 1 Feb 2017 2:27

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tot Ion:330 Resp: 654
 Ion Ratio Lower Upper
 330 100
 332 93.3 75.4 113.0
 141 41.7 31.0 46.6




Manual Integrations
 APPROVED

mohammad
 2/1/2017 2:29:24 PM



#14
 2-Fluorobiphenyl
 Concen: 0.39 ng
 RT: 13.40 min Scan# 881
 Delta R.T. -0.02 min
 Lab File: BE092705.D
 Acq: 1 Feb 2017 2:27

Tgt Ion:172 Resp: 5877
 Ion Ratio Lower Upper
 172 100
 171 38.1 30.1 45.1
 170 28.2 21.8 32.6









#15
Acenaphthylene
Concen: 0.41 ng
RT: 14.49 min Scan# 961
Delta R.T. -0.02 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
152	100		
151	5.0	3.9	5.9
153	12.8	10.4	15.6




Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM



#16
Acenaphthene
Concen: 0.40 ng
RT: 14.85 min Scan# 982
Delta R.T. 0.00 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Tgt Ion	Ratio	Lower	Upper
154	100		
153	440.5	350.8	526.2
152	216.1	171.1	256.7









#17
Fluorene
 Concen: 0.39 ng
 RT: 15.83 min Scan# 1047
 Delta R.T. -0.01 min
 Lab File: BE092705.D
 Acq: 1 Feb 2017 2:27

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
166	100		
165	99.4	78.2	117.4
167	13.3	11.0	16.4




Manual Integrations
 APPROVED

mohammad
 2/1/2017 2:29:24 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.53 min Scan# 1152
 Delta R.T. -0.00 min
 Lab File: BE092705.D
 Acq: 1 Feb 2017 2:27

Tgt Ion	Ratio	Lower	Upper
188	100		
94	8.0	3.2	4.8#
80	7.4	2.6	3.8#





#19

Hexachlorobenzene

Concen: 0.36 ng/ml

RT: 16.87 min Scan# 1123

Delta R.T. 0.00 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion:	284	Resp:	1907
Ion Ratio	Lower	Upper	
284	100		
142	0.0	29.1	43.7#
249	0.0	27.9	41.9#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM



#20

Pentachlorophenol

Concen: 0.41 ng/ml




RT: 17.23 min Scan# 1139


Delta R.T. 0.00 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Tgt Ion:	266	Resp:	547
Ion Ratio	Lower	Upper	
266	100		
268	75.7	62.5	93.7
264	68.8	57.2	85.8









#21
Phenanthrene
Concen: 0.40 ng
RT: 17.58 min Scan# 1154
Delta R.T. -0.00 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
178	100		
176	18.5	15.8	23.8
179	16.4	16.0	24.0




Manual Integrations
APPROVED


mohammad
2/1/2017 2:29:24 PM



#22
Anthracene
Concen: 0.42 ng
RT: 17.67 min Scan# 1158
Delta R.T. -0.02 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.3	15.0	22.4
179	16.0	12.2	18.2





#23

Fluoranthene

Concen: 0.41 ng

RT: 19.58 min Scan# 1264

Delta R.T. -0.02 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion:	202	Resp:	51911
Ion Ratio	Lower	Upper	
202	100		
101	2.8	2.2	3.4
203	17.9	13.7	20.5

Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.72 min Scan# 1374


Delta R.T. -0.00 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Tgt Ion:	240	Resp:	122162
Ion Ratio	Lower	Upper	
240	100		
120	0.8	2.6	4.0#
236	21.2	24.6	37.0#









#25
Pvrene
Concen: 0.46 ng
RT: 19.93 min Scan# 1285
Delta R.T. -0.02 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
202	100		
200	5.3	4.2	6.4
203	18.0	13.9	20.9




Manual Integrations
APPROVED


mohammad
2/1/2017 2:29:24 PM



#26
Terphenyl-d14
Concen: 0.44 ng
RT: 20.16 min Scan# 1298
Delta R.T. -0.00 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.3	6.7	10.1
122	4.8	3.6	5.4





#27

Benzo(a)anthracene

Concen: 0.43 ng

RT: 21.70 min Scan# 1373

Delta R.T. -0.00 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion:	228	Resp:	51700
Ion Ratio	Lower	Upper	
228	100		
226	18.6	23.6	35.4#
229	28.7	13.3	19.9#

Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM



#28

Chrysene

Concen: 0.42 ng m




RT: 21.75 min Scan# 1375


Delta R.T. -0.00 min

Lab File: BE092705.D

Acq: 1 Feb 2017 2:27

Tgt Ion:	228	Resp:	57035
Ion Ratio	Lower	Upper	
228	100		
226	25.1	36.3	54.5#
229	21.6	10.6	16.0#









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 0.74 ng
 RT: 21.61 min Scan# 1369
 Delta R.T. -0.00 min
 Lab File: BE092705.D
 Acq: 1 Feb 2017 2:27

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tot Ion	Ratio	Resp	Lower	Upper
149	100			
167	24.5	16.0	24.0#	
279	0.0	16.0	24.0#	




Manual Integrations
 APPROVED


mohammad
 2/1/2017 2:29:24 PM



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 0.40 ng
 RT: 26.53 min Scan# 1712
 Delta R.T. -0.02 min
 Lab File: BE092705.D
 Acq: 1 Feb 2017 2:27

Tgt Ion	Ratio	Resp	Lower	Upper
276	100			
138	28.5	20.3	30.5	
277	24.9	19.7	29.5	









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.07 min Scan# 1542
Delta R.T. 0.00 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion:	264	Resp:	129450
Ion Ratio	Lower	Upper	
264	100		
260	24.9	20.1	30.1
265	22.1	17.9	26.9




Manual Integrations
APPROVED


mohammad
2/1/2017 2:29:24 PM



#32
Benzo(b)fluoranthene
Concen: 0.42 ng
RT: 23.34 min Scan# 1479
Delta R.T. -0.01 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Tgt Ion:	252	Resp:	13168
Ion Ratio	Lower	Upper	
252	100		
253	24.5	18.7	28.1
125	13.5	10.5	15.7









#33
Benzo(k)fluoranthene
Concen: 0.42 ng
RT: 23.38 min Scan# 1483
Delta R.T. -0.01 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.8	20.3	30.5
125	13.5	9.6	14.4




Manual Integrations
APPROVED


mohammad
2/1/2017 2:29:24 PM



#34
Benzo(a)pyrene
Concen: 0.43 ng
RT: 23.95 min Scan# 1532
Delta R.T. -0.01 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.7	19.0	28.6
125	16.9	11.8	17.8









#35
Dibenzo(a,h)anthracene
Concen: 0.42 ng
RT: 26.55 min Scan# 1713
Delta R.T. -0.03 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
278	100		
139	20.9	13.8	20.8#
279	25.2	21.4	32.2

Manual Integrations
APPROVED

mohammad
2/1/2017 2:29:24 PM



#36
Benzo(g,h,i)perylene
Concen: 0.41 ng
RT: 27.30 min Scan# 1757
Delta R.T. -0.02 min
Lab File: BE092705.D
Acq: 1 Feb 2017 2:27

Tgt Ion	Ratio	Lower	Upper
276	100		
277	26.1	19.8	29.8
138	23.7	19.8	29.6

