

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013118\
 Data File : BE095327.D
 Acq On : 31 Jan 2018 21:09
 Operator : SJ/JU
 Sample : J1245-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6B65

Manual Integrations
 APPROVED

Sohil
 2/1/2018 5:58:32 PM

Quant Time: Feb 01 04:28:13 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 04:05:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	2192	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	7042	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	4704	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	9469m	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	9567	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	8989m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	3982	0.35	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	10578	0.29	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	11.31	128	1840	0.093	ng/ul#	88
5) 2-Methylnaphthalene	12.89	142	1622	0.123	ng/ul	99
8) Acenaphthene	15.08	153	1370	0.078	ng/ul#	93
9) Fluorene	16.03	166	1431m	0.079	ng/ul	
11) Pentachlorophenol	17.36	266	10650	5.807	ng/ul	95
12) Phenanthrene	17.75	178	6827	0.214	ng/ul#	88
13) Anthracene	17.85	178	1109	0.038	ng/ul#	84
15) Fluoranthene	19.72	202	2130	0.052	ng/ul	85
17) Pyrene	20.08	202	1260	0.039	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013118\
Data File : BE095327.D
Acq On : 31 Jan 2018 21:09
Operator : SJ/JU
Sample : J1245-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6B65

Manual Integrations
APPROVED

Sohil
2/1/2018 5:58:32 PM

Quant Time: Feb 01 04:28:13 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 04:05:22 2018
Response via : Initial Calibration

