

Quantitation Report (QT Reviewed)


Data Path : Z:\HPCHEM1\BNA E\DATA\BE013118\
 Data File : BE095324.D
 Acq On : 31 Jan 2018 19:21
 Operator : SJ/JU
 Sample : PB105941BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK41

Manual Integrations
 APPROVED

Sohil
 2/1/2018 5:58:24 PM

Quant Time: Feb 01 04:15:06 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 04:05:22 2018
 Response via : Initial Calibration



Quantitation Report (Qedit)


Data Path : Z:\HPCHEM1\BNA E\DATA\BE013118\
 Data File : BE095324.D
 Acq On : 31 Jan 2018 19:21
 Operator : SJ/JU
 Sample : PB105941BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK41

Manual Integrations
 APPROVED

Sohil
 2/1/2018 5:58:24 PM

Quant Time: Feb 01 04:06:10 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 04:05:22 2018
 Response via : Initial Calibration



(10) Phenanthrene-d10 (I)
 17.706min (-17.706) 0.00ng/ul
 response 0

Ion	Exp%	Act%
188.00	100	0.00
94.00	11.00	0.00#
80.00	11.80	0.00#
0.00	0.00	0.00

Quantitation Report (Qedit)


Data Path : Z:\HPCHEM1\BNA E\DATA\BE013118\
 Data File : BE095324.D
 Acq On : 31 Jan 2018 19:21
 Operator : SJ/JU
 Sample : PB105941BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK41

Manual Integrations
 APPROVED

Sohil
 2/1/2018 5:58:24 PM

Quant Time: Feb 01 04:06:10 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 04:05:22 2018
 Response via : Initial Calibration



(10) Phenanthrene-d10 (I)

17.707min (+0.000) 0.40ng/ul m> SJ

2/9/18

response 8893

Ion	Exp%	Act%
188.00	100	100
94.00	11.00	17.20#
80.00	11.80	18.10#
0.00	0.00	0.00

Quantitation Report (Qedit)


Data Path : Z:\HPCHEM1\BNA E\DATA\BE013118\
 Data File : BE095324.D
 Acq On : 31 Jan 2018 19:21
 Operator : SJ/JU
 Sample : PB105941BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK41

Manual Integrations
 APPROVED

Sohil
 2/1/2018 5:58:24 PM

Quant Time: Feb 01 04:14:24 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 04:05:22 2018
 Response via : Initial Calibration



(20) Perylene-d12 (I)

24.274min (-0.162) 0.40ng/ul

response 569005

Ion	Exp%	Act%
264.00	100	100
260.00	30.10	21.22#
265.00	103.50	21.45#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\HPCHEM1\BNA E\DATA\BE013118\
 Data File : BE095324.D
 Acq On : 31 Jan 2018 19:21
 Operator : SJ/JU
 Sample : PB105941BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK41

Manual Integrations
 APPROVED

Sohil
 2/1/2018 5:58:24 PM

Quant Time: Feb 01 04:14:24 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 04:05:22 2018
 Response via : Initial Calibration



TIC: BE095324.D

(20) Perylene-d12 (I)

24.436min (-0.000) 0.40ng/ul m > SJ 2/9/18

response 8609

Ion	Exp%	Act%
264.00	100	100
260.00	30.10	25.52
265.00	103.50	26.88#
0.00	0.00	0.00

Quantitation Report (QT Reviewed)

Instrument :
BNA_E
ClientSampleId :
SBLK41

Manual Integrations
APPROVED

Sohil
2/1/2018 5:58:24 PM

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013118\
Data File : BE095324.D
Acq On : 31 Jan 2018 19:21
Operator : SJ/JU
Sample : PB105941BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Feb 01 04:15:06 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 04:05:22 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	2220	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	6797	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	4401	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	8893m	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	8906	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	8609m	0.40	ng/ul	0.00

} SJ 2/9/18

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.81	152	4056	0.37	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	10644	0.31	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed