

Data Path : Z:\HPCHEM1\BNA E\DATA\BE013118\
 Data File : BE095355.D
 Acq On : 1 Feb 2018 13:51
 Operator : SJ/JU
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.430

Quant Time: Feb 01 14:26:02 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 06:27:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	76	0.00
3	Naphthalene	1.120	1.119	0.1	76	0.00
4 SURR	2-Methylnaphthalene-d10	0.651	0.691	-6.1	80	0.00
5	2-Methylnaphthalene	0.748	0.783	-4.7	80	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	81	0.00
7	Acenaphthylene	1.925	2.081	-8.1	87	0.00
8 C	Acenaphthene	1.493	1.499	-0.4	81	0.00
9	Fluorene	1.540	1.623	-5.4	85	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	73	0.00
11	Pentachlorophenol	0.077	0.084	-9.1	83	0.00
12	Phenanthrene	1.345	1.345	0.0	71	0.00
13	Anthracene	1.231	1.269	-3.1	74	0.00
14 SURR	Fluoranthene-d10	1.531	1.401	8.5	68	0.00
15 C	Fluoranthene	1.742	1.557	10.6	66	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
17	Pyrene	1.342	1.558	-16.1	73	0.00
18	Benzo(a)anthracene	1.251	1.361	-8.8	65	0.00
19	Chrysene	1.299	1.296	0.2	59	0.00
20 I	Perylene-d12	1.000	1.000	0.0	66	0.00
21	Benzo(b)fluoranthene	1.411	1.319	6.5	61	0.00
22	Benzo(k)fluoranthene	1.291	1.281	0.8	65	0.00
23 C	Benzo(a)pyrene	1.260	1.252	0.6	65	0.00
24	Indeno(1,2,3-cd)pyrene	1.382	1.407	-1.8	67	0.00
25	Dibenzo(a,h)anthracene	1.130	1.138	-0.7	65	0.00
26	Benzo(a,h,i)perylene	1.170	1.181	-0.9	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0