

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020117\
 Data File : BE092710.D
 Acq On : 1 Feb 2017 11:51
 Operator : SJ/MA
 Sample : I1505-07DL 10X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SB-01BDL

Quant Time: Feb 01 14:26:01 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	10323	5.00	ng	0.00
7) Naphthalene-d8	10.91	136	47150	5.00	ng	0.00
12) Acenaphthene-d10	14.78	164	29053	5.00	ng	0.00
18) Phenanthrene-d10	17.53	188	61527	5.00	ng	0.00
24) Chrysene-d12	21.72	240	64230	5.00	ng	0.00
31) Perylene-d12	24.06	264	72859	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.64	112	1304	0.57	ng	-0.01
5) Phenol-d6	7.31	99	1027	0.36	ng	-0.02
8) Nitrobenzene-d5	9.34	82	1365	0.45	ng	0.00
13) 2,4,6-Tribromophenol	16.28	330	717	0.95	ng	-0.01
14) 2-Fluorobiphenyl	13.41	172	3767	0.53	ng	-0.01
26) Terphenyl-d14	20.16	244	5497	0.61	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.95	128	38972	3.96	ng	# 95
11) 2-Methylnaphthalene	12.60	142	3509	0.57	ng	98
16) Acenaphthene	14.85	154	2974	0.46	ng	95
17) Fluorene	15.84	166	1959	0.24	ng	96
21) Phenanthrene	17.58	178	3605	0.27	ng	96
23) Fluoranthene	19.59	202	2636	0.04	ng	99
25) Pyrene	19.95	202	1992	0.03	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020117\
Data File : BE092710.D
Acq On : 1 Feb 2017 11:51
Operator : SJ/MA
Sample : I1505-07DL 10X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SB-01BDL

Quant Time: Feb 01 14:26:01 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 17 10:14:01 2017
Response via : Initial Calibration

