

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020117\
 Data File : BE092716.D
 Acq On : 1 Feb 2017 15:44
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 02 09:53:01 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	154	0.00
2	1,4-Dioxane	0.400	0.385	3.8	143	0.00
3	n-Nitrosodimethylamine	0.400	0.344	14.0	152	0.00
4 S	2-Fluorophenol	0.400	0.430	-7.5	195	-0.01
5 S	Phenol-d6	0.400	0.479	-19.7	225	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.363	9.3	167	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	161	0.00
8 S	Nitrobenzene-d5	0.400	0.386	3.5	187	-0.02
9	Naphthalene	0.400	0.402	-0.5	175	-0.02
10	Hexachlorobutadiene	0.400	0.431	-7.7	183	-0.02
11	2-Methylnaphthalene	0.400	0.402	-0.5	181	-0.02
12 I	Acenaphthene-d10	5.000	5.000	0.0	156	0.00
13 S	2,4,6-Tribromophenol	0.400	0.506	-26.5#	219	-0.01
14 S	2-Fluorobiphenyl	0.400	0.412	-3.0	173	-0.02
15	Acenaphthylene	0.400	0.421	-5.2	185	-0.02
16	Acenaphthene	0.400	0.430	-7.5	177	-0.02
17	Fluorene	0.400	0.409	-2.2	178	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	158	0.00
19	Hexachlorobenzene	0.400	0.327	18.3	135	0.00
20	Pentachlorophenol	0.400	0.488	-22.0	211	-0.02
21	Phenanthrene	0.400	0.379	5.3	155	0.00
22	Anthracene	0.400	0.432	-8.0	192	-0.02
23	Fluoranthene	0.400	0.445	-11.2	194	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	133	0.00
25	Pyrene	0.400	0.482	-20.5	191	-0.02
26 S	Terphenyl-d14	0.400	0.461	-15.3	178	0.00
27	Benzo(a)anthracene	0.400	0.471	-17.7	181	0.00
28	Chrysene	0.400	0.414	-3.5	147	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.816	-104.0#	313	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.429	-7.2	162	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	138	0.00
32	Benzo(b)fluoranthene	0.400	0.422	-5.5	158	-0.01
33	Benzo(k)fluoranthene	0.400	0.425	-6.2	171	-0.01
34 C	Benzo(a)pyrene	0.400	0.428	-7.0	171	-0.01
35	Dibenzo(a,h)anthracene	0.400	0.422	-5.5	164	-0.02
36	Benzo(g,h,i)perylene	0.400	0.419	-4.7	160	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0