

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020117\
 Data File : BE092716.D
 Acq On : 1 Feb 2017 15:44
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 02 09:53:01 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE011617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 10:14:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	154#	0.00
2	1,4-Dioxane	0.775	0.756	2.5	143	0.00
3	n-Nitrosodimethylamine	0.902	0.777	13.9	152#	0.00
4 S	2-Fluorophenol	1.116	1.201	-7.6	195#	-0.01
5 S	Phenol-d6	1.368	1.638	-19.7	225#	-0.02
6	bis(2-Chloroethyl)ether	1.453	1.320	9.2	167#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	161#	0.00
8 S	Nitrobenzene-d5	0.325	0.313	3.7	187#	-0.02
9	Naphthalene	1.044	1.050	-0.6	175#	-0.02
10	Hexachlorobutadiene	0.154	0.166	-7.8	183#	-0.02
11	2-Methylnaphthalene	0.648	0.652	-0.6	181#	-0.02
12 I	Acenaphthene-d10	1.000	1.000	0.0	156#	0.00
13 S	2,4,6-Tribromophenol	0.126	0.141	-11.9	219#	-0.01
14 S	2-Fluorobiphenyl	1.230	1.266	-2.9	173#	-0.02
15	Acenaphthylene	6.890	7.243	-5.1	185#	-0.02
16	Acenaphthene	1.105	1.189	-7.6	177#	-0.02
17	Fluorene	1.387	1.416	-2.1	178#	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	158#	0.00
19	Hexachlorobenzene	0.189	0.169	10.6	135	0.00
20	Pentachlorophenol	0.051	0.063	-23.5	211#	-0.02
21	Phenanthrene	1.104	1.047	5.2	155#	0.00
22	Anthracene	1.018	1.099	-8.0	192#	-0.02
23	Fluoranthene	4.895	5.446	-11.3	194#	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
25	Pyrene	4.929	5.941	-20.5	191#	-0.02
26 S	Terphenyl-d14	0.707	0.814	-15.1	178#	0.00
27	Benzo(a)anthracene	4.926	5.798	-17.7	181#	0.00
28	Chrysene	5.542	5.732	-3.4	147	0.00
29	Bis(2-ethylhexyl)phthalate	0.913	1.542	-68.9#	313#	0.00
30	Indeno(1,2,3-cd)pyrene	5.987	6.417	-7.2	162#	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	138	0.00
32	Benzo(b)fluoranthene	1.210	1.275	-5.4	158#	-0.01
33	Benzo(k)fluoranthene	1.215	1.291	-6.3	171#	-0.01
34 C	Benzo(a)pyrene	1.136	1.215	-7.0	171#	-0.01
35	Dibenzo(a,h)anthracene	1.112	1.173	-5.5	164#	-0.02
36	Benzo(g,h,i)perylene	4.699	4.926	-4.8	160#	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0