

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020118\  
 Data File : BE095367.D  
 Acq On : 1 Feb 2018 21:06  
 Operator : SJ/JU  
 Sample : J1282-02  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :

Quant Time: Feb 02 04:09:04 2018  
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 02 04:07:20 2018  
 Response via : Initial Calibration


| Internal Standards          | R.T.  | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|-------|--------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.44  | 152  | 2125     | 0.40  | ng/ul  | 0.00     |
| 2) Naphthalene-d8           | 11.26 | 136  | 6426     | 0.40  | ng/ul  | 0.00     |
| 6) Acenaphthene-d10         | 15.02 | 164  | 4203     | 0.40  | ng/ul  | 0.00     |
| 10) Phenanthrene-d10        | 0.00  | 188  | 0        | 0.00  | ng/ul  | -17.71   |
| 16) Chrysene-d12            | 21.80 | 240  | 8642     | 0.40  | ng/ul  | 0.00     |
| 20) Perylene-d12            | 24.27 | 264  | 544102   | 0.40  | ng/ul  | -0.16    |
| System Monitoring Compounds |       |      |          |       |        |          |
| 4) 2-Methylnaphthalene-d10  | 12.81 | 152  | 3964     | 0.38  | ng/ul  | 0.00     |
| 14) Fluoranthene-d10        | 19.69 | 212  | 10802    | 0.00  | ng/ul  | 0.00     |
| Target Compounds            |       |      |          |       |        |          |
| 17) Pyrene                  | 20.05 | 202  | 1855     | 0.064 | ng/ul# | Ovalue 1 |




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020118\  
Data File : BE095367.D  
Acq On : 1 Feb 2018 21:06  
Operator : SJ/JU  
Sample : J1282-02  
Misc :  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :

Quant Time: Feb 02 04:09:04 2018  
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Feb 02 04:07:20 2018  
Response via : Initial Calibration





#17

Pvrene

Concen: 0.064 ng/ul

RT: 20.05 min Scan# 2460

Delta R.T. -0.03 min

Lab File: BE095367.D

Acq: 1 Feb 2018 21:06

Instrument :

BNA\_E

ClientSampleId :

Tot Ion:202 Resp: 1855

Ion Ratio Lower Upper

202 100

101 165.9 18.2 27.4#

100 1064.6 15.6 23.4#

