

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020118\
 Data File : BE095364.D
 Acq On : 1 Feb 2018 19:19
 Operator : SJ/JU
 Sample : J1282-09
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6B90

Manual Integrations
 APPROVED

Quant Time: Feb 02 04:29:43 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 02 04:07:20 2018
 Response via : Initial Calibration

Soniii

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)	2/2/2018 4:41:51 PM
1) 1,4-Dichlorobenzene-d4	8.44	152	1953	0.40	ng/ul	0.00	
2) Naphthalene-d8	11.26	136	5971	0.40	ng/ul	0.00	
6) Acenaphthene-d10	15.02	164	3972	0.40	ng/ul	0.00	
10) Phenanthrene-d10	17.71	188	7848m	0.40	ng/ul	0.00	
16) Chrysene-d12	21.80	240	7571	0.40	ng/ul	0.00	
20) Perylene-d12	24.44	264	7299m	0.40	ng/ul	0.00	
System Monitoring Compounds							
4) 2-Methylnaphthalene-d10	12.81	152	4710	0.48	ng/ul	0.00	
14) Fluoranthene-d10	19.69	212	12015	0.40	ng/ul	0.00	
Target Compounds							
12) Phenanthrene	17.75	178	874m	0.033	ng/ul	Ovalue	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020118\
Data File : BE095364.D
Acq On : 1 Feb 2018 19:19
Operator : SJ/JU
Sample : J1282-09
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6B90

Manual Integrations
APPROVED

Quant Time: Feb 02 04:29:43 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 02 04:07:20 2018
Response via : Initial Calibration

Sum

