

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020118\
 Data File : BE095382.D
 Acq On : 2 Feb 2018 6:04
 Operator : SJ/JU
 Sample : J1295-13
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6C11

Manual Integrations
 APPROVED

Sohil
 2/2/2018 4:42:39 PM

Quant Time: Feb 02 07:45:12 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 02 04:53:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	2500	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	8101	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	5134	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	10552m	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	10050	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	8648m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	4529	0.34	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	10968	0.27	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	11.31	128	1462	0.064	ng/ul#	74
5) 2-Methylnaphthalene	12.89	142	610	0.040	ng/ul	99
7) Acenaphthylene	14.74	152	4006	0.162	ng/ul#	89
11) Pentachlorophenol	17.36	266	404982	198.152	ng/ul#	56
12) Phenanthrene	17.75	178	45243	1.275	ng/ul#	88
13) Anthracene	17.85	178	8399	0.259	ng/ul	94
15) Fluoranthene	19.72	202	97841	2.129	ng/ul	83
17) Pyrene	20.08	202	75881	2.250	ng/ul#	78
18) Benzo(a)anthracene	21.79	228	15927	0.507	ng/ul#	90
19) Chrysene	21.84	228	37917	1.162	ng/ul	97
21) Benzo(b)fluoranthene	23.62	252	47548m	1.558	ng/ul	
22) Benzo(k)fluoranthene	23.67	252	9905m	0.355	ng/ul	
23) Benzo(a)pyrene	24.32	252	4917m	0.180	ng/ul	
24) Indeno(1,2,3-cd)pyrene	27.27	276	12471m	0.417	ng/ul	
25) Dibenzo(a,h)anthracene	27.27	278	3163m	0.129	ng/ul	
26) Benzo(a,h,i)perylene	28.13	276	9891m	0.391	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020118\
Data File : BE095382.D
Acq On : 2 Feb 2018 6:04
Operator : SJ/JU
Sample : J1295-13
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6C11

Manual Integrations
APPROVED

Sohil
2/2/2018 4:42:39 PM

Quant Time: Feb 02 07:45:12 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 02 04:53:51 2018
Response via : Initial Calibration

