

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020118\
 Data File : BE095357.D
 Acq On : 1 Feb 2018 15:06
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.431

Quant Time: Feb 02 04:04:51 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 06:27:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	73	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	77	0.00
3	Naphthalene	0.400	0.401	-0.3	78	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.427	-6.7	82	0.00
5	2-Methylnaphthalene	0.400	0.429	-7.2	83	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	83	0.00
7	Acenaphthylene	0.400	0.431	-7.7	89	0.00
8 C	Acenaphthene	0.400	0.404	-1.0	83	0.00
9	Fluorene	0.400	0.419	-4.7	86	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	75	0.00
11	Pentachlorophenol	0.400	0.358	10.5	70	0.00
12	Phenanthrene	0.400	0.393	1.8	72	0.00
13	Anthracene	0.400	0.411	-2.7	76	0.00
14 SURR	Fluoranthene-d10	0.400	0.357	10.8	68	0.00
15 C	Fluoranthene	0.400	0.351	12.3	67	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	61	0.00
17	Pyrene	0.400	0.477	-19.2	75	0.00
18	Benzo(a)anthracene	0.400	0.434	-8.5	65	0.00
19	Chrysene	0.400	0.396	1.0	59	0.00
20 I	Perylene-d12	0.400	0.400	0.0	66	0.00
21	Benzo(b)fluoranthene	0.400	0.377	5.8	62	0.00
22	Benzo(k)fluoranthene	0.400	0.398	0.5	65	0.00
23 C	Benzo(a)pyrene	0.400	0.400	0.0	66	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.405	-1.3	66	0.00
25	Dibenzo(a,h)anthracene	0.400	0.399	0.3	65	0.00
26	Benzo(a,h,i)perylene	0.400	0.402	-0.5	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0