

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
 Data File : BE089460.D
 Acq On : 2 Feb 2015 17:49
 Operator : TP/IZ
 Sample : SSTDICC0.5
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.5

Quant Time: Feb 03 03:16:19 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 03 03:06:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.58	152	35079	5.00	ng	0.00
7) Naphthalene-d8	12.49	136	126319	5.00	ng	0.00
13) Acenaphthene-d10	16.68	164	45506	5.00	ng	0.00
17) Phenanthrene-d10	19.99	188	77529	5.00	ng	0.00
20) Chrysene-d12	23.84	240	109410	5.00	ng	0.00
27) Perylene-d12	25.56	264	135189	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 2-Fluorophenol	6.75	112	5700	0.47	ng	0.00
4) Phenol-d6	8.87	99	6748	0.46	ng	0.00
8) Nitrobenzene-d5	10.85	82	4444	0.44	ng	0.00
14) 2,4,6-Tribromophenol	18.59	330	478	0.31	ng	0.00
15) 2-Fluorobiphenyl	15.13	172	6314	0.50	ng	0.00
22) Terphenyl-d14	22.49	244	5673	0.41	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) n-Nitrosodimethylamine	3.64	42	2499	0.40	ng	# 51
5) Phenol	8.90	94	7156	0.45	ng	92
6) bis(2-Chloroethyl)ether	9.06	93	5349	0.45	ng	# 96
9) Nitrobenzene	10.89	77	4964	0.42	ng	98
10) 2,4-Dimethylphenol	11.83	122	3596	0.43	ng	99
11) 2,4-Dichlorophenol	12.21	162	2596	0.42	ng	98
12) Hexachlorobutadiene	12.87	225	1995	0.48	ng	99
16) 2,4-Dinitrophenol	16.92	184	59	0.10	ng	# 59
18) Hexachlorobenzene	19.23	284	2068	0.50	ng	97
19) Pentachlorophenol	19.67	266	145	0.20	ng	# 80
21) Benzidine	22.14	184	3080	0.41	ng	# 96
23) Benzo(a)anthracene	23.83	228	39081	0.46	ng	98
24) 3,3'-Dichlorobenzidine	23.82	252	3262	0.51	ng	# 95
25) Chrysene	23.87	228	52934	0.48	ng	98
26) Indeno(1,2,3-cd)pyrene	27.07	276	11519	0.65	ng	# 79
28) Benzo(b)fluoranthene	25.13	252	15418	1.27	ng	97
29) Benzo(k)fluoranthene	25.13	252	15418	0.42	ng	98
30) Benzo(a)pyrene	25.49	252	7975	0.49	ng	# 93
31) Dibenzo(a,h)anthracene	27.10	278	9596	0.60	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
Data File : BE089460.D
Acq On : 2 Feb 2015 17:49
Operator : TP/IZ
Sample : SSTDICC0.5
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.5

Quant Time: Feb 03 03:16:19 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 03 03:06:34 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 9.58 min Scan# 1131

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDIC0.5


Tgt Ion:152 Resp: 35079

Ion Ratio Lower Upper

152 100

150 150.9 124.6 186.8




115 66.2 55.4 83.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

n-Nitrosodimethylamine

Concen: 0.40 ng

RT: 3.64 min Scan# 330

Delta R.T. 0.03 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion: 42 Resp: 2499

Ion Ratio Lower Upper

42 100

74 120.2 72.1 108.1#



44 89.4 28.0 42.0#




Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0





#3

2-Fluorophenol

Concen: 0.47 ng

RT: 6.75 min Scan# 789

Delta R.T. -0.01 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDICC0.5


Tgt Ion: 112 Resp: 5700

Ion Ratio Lower Upper

112 100

64 64.2 49.8 74.8




63 34.6 24.9 37.3



Abundance Ion 112.00 (111.70 to 112.70): BE0

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0



#4

Phenol-d6

Concen: 0.46 ng

RT: 8.87 min Scan# 1051

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion: 99 Resp: 6748

Ion Ratio Lower Upper

99 100

42 15.8 15.0 22.4



71 35.5 30.6 46.0




Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0





#5

Phenol

Concen: 0.45 ng

RT: 8.90 min Scan# 1054

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDICC0.5





Tgt Ion: 94 Resp: 7156

Ion Ratio Lower Upper

94 100

65 113.9 105.0 145.0

66 41.6 22.4 62.4



#6

bis(2-Chloroethyl)ether

Concen: 0.45 ng

RT: 9.06 min Scan# 1072

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49




Tgt Ion: 93 Resp: 5349


Ion Ratio Lower Upper

93 100

63 70.2 59.2 88.8

95 0.0 0.0 0.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 12.49 min Scan# 1562

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDIC0.5


Tgt Ion: 136 Resp: 126319

Ion Ratio Lower Upper

136 100

137 10.9 8.6 13.0




54 7.3 6.5 9.7



Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0



#8

Nitrobenzene-d5

Concen: 0.44 ng

RT: 10.85 min Scan# 1349

Delta R.T. -0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion: 82 Resp: 4444

Ion Ratio Lower Upper

82 100

128 42.6 34.2 51.2



54 49.8 37.3 55.9




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Nitrobenzene

Concen: 0.42 ng

RT: 10.89 min Scan# 1358

Delta R.T. -0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDIC0.5


Tgt Ion: 77 Resp: 4964

Ion Ratio Lower Upper

77 100

123 43.1 33.6 50.4




65 13.3 11.0 16.6



Abundance Ion 77.00 (76.70 to 77.70): BE089460.D (-1249) (-)

Ion 123.00 (122.70 to 123.70): BE089460.D (-1249) (-)

Ion 65.00 (64.70 to 65.70): BE089460.D (-1249) (-)



#10

2,4-Dimethylphenol

Concen: 0.43 ng

RT: 11.83 min Scan# 1494

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion: 122 Resp: 3596

Ion Ratio Lower Upper

122 100

107 113.6 90.6 136.0



121 56.5 43.7 65.5




Abundance Ion 122.00 (121.70 to 122.70): BE089460.D (-1474) (-)

Ion 107.00 (106.70 to 107.70): BE089460.D (-1474) (-)

Ion 121.00 (120.70 to 121.70): BE089460.D (-1474) (-)





#11

2,4-Dichlorophenol

Concen: 0.42 ng

RT: 12.21 min Scan# 1533

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDIC0.5


Tgt Ion:162 Resp: 2596

Ion Ratio Lower Upper

162 100

164 62.2 40.5 80.5




98 49.0 28.1 68.1



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): E



#12

Hexachlorobutadiene

Concen: 0.48 ng

RT: 12.87 min Scan# 1601

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion:225 Resp: 1995

Ion Ratio Lower Upper

225 100

223 63.4 50.2 75.2



227 62.8 49.9 74.9




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E









#13
Acenaphthene-d10
Concen: 5.00 ng
RT: 16.68 min Scan# 2755
Delta R.T. 0.00 min
Lab File: BE089460.D
Acq: 2 Feb 2015 17:49




Instrument :
BNA_E
LabSampleId :
SSTDIC0.5


Tgt Ion:164 Resp: 45506
Ion Ratio Lower Upper
164 100
162 95.7 77.4 116.2
160 43.7 35.5 53.3



#14
2,4,6-Tribromophenol
Concen: 0.31 ng
RT: 18.59 min Scan# 3037
Delta R.T. 0.01 min
Lab File: BE089460.D
Acq: 2 Feb 2015 17:49

Tgt Ion:330 Resp: 478
Ion Ratio Lower Upper
330 100
332 97.9 75.4 113.2
141 46.9 35.7 53.5









#15
2-Fluorobiphenyl
Concen: 0.50 ng
RT: 15.13 min Scan# 2426
Delta R.T. 0.01 min
Lab File: BE089460.D
Acq: 2 Feb 2015 17:49




Instrument :
BNA_E
LabSampleId :
SSTDIC0.5


Tgt Ion:172 Resp: 6314
Ion Ratio Lower Upper
172 100
171 32.7 27.7 41.5
170 20.3 18.3 27.5



#16
2,4-Dinitrophenol
Concen: 0.10 ng
RT: 16.92 min Scan# 2791
Delta R.T. 0.01 min
Lab File: BE089460.D
Acq: 2 Feb 2015 17:49

Tgt Ion:184 Resp: 59
Ion Ratio Lower Upper
184 100
63 136.0 77.0 115.6#
154 101.3 54.0 81.0#





#17

Phenanthrene-d10

Concen: 5.00 ng

RT: 19.99 min Scan# 3240

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDIC0.5


Tgt Ion:188 Resp: 77529

Ion Ratio Lower Upper

188 100

94 12.8 10.2 15.4




80 13.0 10.9 16.3



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0



#18

Hexachlorobenzene

Concen: 0.50 ng

RT: 19.23 min Scan# 3128

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion:284 Resp: 2068

Ion Ratio Lower Upper

284 100

142 42.3 31.8 47.8



249 29.6 23.5 35.3




Abundance Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E





#19

Pentachlorophenol

Concen: 0.20 ng

RT: 19.67 min Scan# 3192

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDICC0.5


Tgt Ion:266 Resp: 145

Ion Ratio Lower Upper




266 100

268 81.1 52.6 78.8#

264 81.1 52.2 78.2#



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E



#20

Chrysene-d12

Concen: 5.00 ng

RT: 23.84 min Scan# 3873

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion:240 Resp: 109410

Ion Ratio Lower Upper



240 100


120 8.4 6.2 9.4

236 22.9 18.2 27.4



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#21

Benzidine

Concen: 0.41 ng

RT: 22.14 min Scan# 3600

Delta R.T. -0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDICC0.5


Tgt Ion:184 Resp: 3080

Ion Ratio Lower Upper




184 100

185 15.8 11.7 17.5

183 13.6 9.4 14.0



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 185.00 (184.70 to 185.70): E
Ion 183.00 (182.70 to 183.70): E



#22

Terphenyl-d14

Concen: 0.41 ng

RT: 22.49 min Scan# 3675

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion:244 Resp: 5673

Ion Ratio Lower Upper



244 100


212 8.0 5.3 7.9#

122 14.2 10.5 15.7



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E





#23

Benzo(a)anthracene

Concen: 0.46 ng

RT: 23.83 min Scan# 3871

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDIC0.5


Tgt Ion:228 Resp: 39081

Ion Ratio Lower Upper

228 100

226 28.6 21.7 32.5

229 18.9 14.8 22.2






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#24

3,3'-Dichlorobenzidine

Concen: 0.51 ng

RT: 23.82 min Scan# 3870

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion:252 Resp: 3262

Ion Ratio Lower Upper

252 100

254 65.7 54.6 82.0

126 15.6 8.0 12.0#






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#25

Chrysene

Concen: 0.48 ng

RT: 23.87 min Scan# 3876

Delta R.T. 0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDIC0.5


Tgt Ion:228 Resp: 52934

Ion Ratio Lower Upper

228 100

226 29.6 23.8 35.8




229 19.4 14.3 21.5



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#26

Indeno(1,2,3-cd)pyrene

Concen: 0.65 ng

RT: 27.07 min Scan# 4516

Delta R.T. -0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion:276 Resp: 11519

Ion Ratio Lower Upper

276 100

138 9.4 23.3 34.9#



277 24.4 19.0 28.4




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#27
Perylene-d12
 Concen: 5.00 ng
 RT: 25.56 min Scan# 4187
 Delta R.T. -0.00 min
 Lab File: BE089460.D
 Acq: 2 Feb 2015 17:49




Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.5


Tgt Ion:264 Resp: 135189
 Ion Ratio Lower Upper
 264 100
 260 23.8 18.7 28.1
 265 21.4 17.0 25.6



#28
Benzo(b)fluoranthene
 Concen: 1.27 ng
 RT: 25.13 min Scan# 4094
 Delta R.T. 0.03 min
 Lab File: BE089460.D
 Acq: 2 Feb 2015 17:49

Tgt Ion:252 Resp: 15418
 Ion Ratio Lower Upper
 252 100
 253 22.2 17.4 26.0
 125 15.0 10.4 15.6





#29

Benzo(k)fluoranthene

Concen: 0.42 ng

RT: 25.13 min Scan# 4094

Delta R.T. -0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDIC0.5


Tgt Ion:252 Resp: 15418

Ion Ratio Lower Upper

252 100

253 22.2 17.5 26.3




125 15.0 10.3 15.5



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#30

Benzo(a)pyrene

Concen: 0.49 ng

RT: 25.49 min Scan# 4172

Delta R.T. -0.00 min

Lab File: BE089460.D

Acq: 2 Feb 2015 17:49


Tgt Ion:252 Resp: 7975

Ion Ratio Lower Upper

252 100

253 24.7 17.5 26.3



125 18.7 11.8 17.8#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#31

Dibenzo(a,h)anthracene

Concen: 0.60 ng

RT: 27.10 min Scan# 4522

Delta R.T. -0.00 min

Lab File: BE089460.D


Acq: 2 Feb 2015 17:49

Instrument :

BNA_E

LabSampleId :

SSTDIC0.5




Tgt Ion: 278 Resp: 9596

Ion Ratio Lower Upper

278 100

139 24.9 17.1 25.7

279 26.4 18.8 28.2



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E

