

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
 Data File : BE089458.D
 Acq On : 2 Feb 2015 16:31
 Operator : TP/IZ
 Sample : SSTDICCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICCC001

Quant Time: Feb 03 03:15:31 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 03 03:06:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.58	152	43906	5.00	ng	0.00
7) Naphthalene-d8	12.49	136	166168	5.00	ng	0.00
13) Acenaphthene-d10	16.68	164	67628	5.00	ng	0.00
17) Phenanthrene-d10	19.99	188	132399	5.00	ng	0.00
20) Chrysene-d12	23.84	240	168407	5.00	ng	0.00
27) Perylene-d12	25.56	264	180643	5.00	ng	0.00

System Monitoring Compounds						
3) 2-Fluorophenol	6.76	112	14110	0.92	ng	0.00
4) Phenol-d6	8.87	99	16977	0.92	ng	0.00
8) Nitrobenzene-d5	10.85	82	11794	0.88	ng	0.00
14) 2,4,6-Tribromophenol	18.59	330	1774	0.78	ng	0.00
15) 2-Fluorobiphenyl	15.13	172	17940	0.96	ng	0.00
22) Terphenyl-d14	22.48	244	20860	0.99	ng	0.00

Target Compounds					Qvalue
2) n-Nitrosodimethylamine	3.61	42	7437	0.94	ng 100
5) Phenol	8.90	94	18430	0.92	ng 100
6) bis(2-Chloroethyl)ether	9.06	93	13200	0.89	ng # 100
9) Nitrobenzene	10.89	77	13760	0.89	ng 100
10) 2,4-Dimethylphenol	11.83	122	9801	0.89	ng 100
11) 2,4-Dichlorophenol	12.21	162	7407	0.92	ng 100
12) Hexachlorobutadiene	12.87	225	5033	0.92	ng 100
16) 2,4-Dinitrophenol	16.91	184	423	0.49	ng 100
18) Hexachlorobenzene	19.23	284	6595	0.93	ng 100
19) Pentachlorophenol	19.67	266	639	0.52	ng 100
21) Benzidine	22.15	184	13122	1.14	ng 100
23) Benzo(a)anthracene	23.83	228	138201	1.05	ng 100
24) 3,3'-Dichlorobenzidine	23.82	252	11902	1.21	ng 100
25) Chrysene	23.87	228	159780	0.93	ng 100
26) Indeno(1,2,3-cd)pyrene	27.07	276	43530	1.60	ng 100
28) Benzo(b)fluoranthene	25.10	252	25746	1.58	ng 100
29) Benzo(k)fluoranthene	25.13	252	48146	0.98	ng 100
30) Benzo(a)pyrene	25.49	252	29000	1.32	ng 100
31) Dibenzo(a,h)anthracene	27.10	278	35094	1.63	ng 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
Data File : BE089458.D
Acq On : 2 Feb 2015 16:31
Operator : TP/IZ
Sample : SSTDICCC001
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICCC001

Quant Time: Feb 03 03:15:31 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 03 03:06:34 2015
Response via : Initial Calibration

