

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
 Data File : BE089459.D
 Acq On : 2 Feb 2015 17:10
 Operator : TP/IZ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.8

Quant Time: Feb 03 03:15:55 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 03 03:06:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.58	152	41137	5.00	ng	0.00
7) Naphthalene-d8	12.49	136	155410	5.00	ng	0.00
13) Acenaphthene-d10	16.68	164	59516	5.00	ng	0.00
17) Phenanthrene-d10	19.99	188	110341	5.00	ng	0.00
20) Chrysene-d12	23.84	240	142598	5.00	ng	0.00
27) Perylene-d12	25.56	264	160045	5.00	ng	0.00
System Monitoring Compounds						
3) 2-Fluorophenol	6.75	112	10467	0.73	ng	0.00
4) Phenol-d6	8.87	99	12645	0.73	ng	0.00
8) Nitrobenzene-d5	10.85	82	8598	0.69	ng	0.00
14) 2,4,6-Tribromophenol	18.58	330	1080	0.54	ng	0.00
15) 2-Fluorobiphenyl	15.13	172	12559	0.76	ng	0.00
22) Terphenyl-d14	22.48	244	13195	0.74	ng	0.00
Target Compounds						Qvalue
2) n-Nitrosodimethylamine	3.61	42	5392	0.73	ng	# 80
5) Phenol	8.90	94	13435	0.71	ng	95
6) bis(2-Chloroethyl)ether	9.06	93	9774	0.70	ng	# 97
9) Nitrobenzene	10.89	77	9902	0.69	ng	100
10) 2,4-Dimethylphenol	11.83	122	7082	0.69	ng	97
11) 2,4-Dichlorophenol	12.21	162	5211	0.69	ng	98
12) Hexachlorobutadiene	12.87	225	3802	0.75	ng	98
16) 2,4-Dinitrophenol	16.91	184	223	0.29	ng	94
18) Hexachlorobenzene	19.23	284	4420	0.75	ng	98
19) Pentachlorophenol	19.67	266	357	0.35	ng	97
21) Benzidine	22.15	184	7272	0.74	ng	97
23) Benzo(a)anthracene	23.83	228	87216	0.78	ng	99
24) 3,3'-Dichlorobenzidine	23.82	252	7165	0.86	ng	96
25) Chrysene	23.87	228	107032	0.74	ng	99
26) Indeno(1,2,3-cd)pyrene	27.07	276	28363	1.23	ng	# 78
28) Benzo(b)fluoranthene	25.10	252	14807	1.03	ng	99
29) Benzo(k)fluoranthene	25.13	252	33778	0.78	ng	99
30) Benzo(a)pyrene	25.48	252	17803	0.92	ng	98
31) Dibenzo(a,h)anthracene	27.10	278	21421	1.12	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
Data File : BE089459.D
Acq On : 2 Feb 2015 17:10
Operator : TP/IZ
Sample : SSTDICC0.8
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.8

Quant Time: Feb 03 03:15:55 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 03 03:06:34 2015
Response via : Initial Calibration

