

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
 Data File : BE089460.D
 Acq On : 2 Feb 2015 17:49
 Operator : TP/IZ
 Sample : SSTDICC0.5
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.5

Quant Time: Feb 03 03:23:43 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 03 03:06:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.58	152	35079	5.00	ng	0.00
7) Naphthalene-d8	12.49	136	126319	5.00	ng	0.00
13) Acenaphthene-d10	16.68	164	45506	5.00	ng	0.00
17) Phenanthrene-d10	19.99	188	77529	5.00	ng	0.00
20) Chrysene-d12	23.84	240	109410	5.00	ng	0.00
27) Perylene-d12	25.56	264	135189	5.00	ng	0.00

System Monitoring Compounds

3) 2-Fluorophenol	6.75	112	5700	0.47	ng	0.00
4) Phenol-d6	8.87	99	6748	0.46	ng	0.00
8) Nitrobenzene-d5	10.85	82	4444	0.44	ng	0.00
14) 2,4,6-Tribromophenol	18.59	330	478	0.31	ng	0.00
15) 2-Fluorobiphenyl	15.13	172	6314	0.50	ng	0.00
22) Terphenyl-d14	22.49	244	5673	0.41	ng	0.00

Target Compounds

						Qvalue
2) n-Nitrosodimethylamine	3.64	42	2499	0.40	ng	# 51
5) Phenol	8.90	94	7156	0.45	ng	92
6) bis(2-Chloroethyl)ether	9.06	93	5349	0.45	ng	# 96
9) Nitrobenzene	10.89	77	4964	0.42	ng	98
10) 2,4-Dimethylphenol	11.83	122	3596	0.43	ng	99
11) 2,4-Dichlorophenol	12.21	162	2596	0.42	ng	98
12) Hexachlorobutadiene	12.87	225	1995	0.48	ng	99
16) 2,4-Dinitrophenol	16.92	184	59	0.10	ng	# 59
18) Hexachlorobenzene	19.23	284	2068	0.50	ng	97
19) Pentachlorophenol	19.67	266	145	0.20	ng	# 80
21) Benzidine	22.14	184	3080	0.41	ng	96
23) Benzo(a)anthracene	23.83	228	39081	0.46	ng	98
24) 3,3'-Dichlorobenzidine	23.82	252	3262	0.51	ng	# 95
25) Chrysene	23.87	228	52934	0.48	ng	98
26) Indeno(1,2,3-cd)pyrene	27.07	276	13821m	0.78	ng	
28) Benzo(b)fluoranthene	25.10	252	5980m	0.49	ng	
29) Benzo(k)fluoranthene	25.13	252	17179m	0.47	ng	
30) Benzo(a)pyrene	25.49	252	7975	0.49	ng	# 93
31) Dibenzo(a,h)anthracene	27.10	278	9596	0.60	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
Data File : BE089460.D
Acq On : 2 Feb 2015 17:49
Operator : TP/IZ
Sample : SSTDICC0.5
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.5

Quant Time: Feb 03 03:23:43 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 03 03:06:34 2015
Response via : Initial Calibration

