

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
 Data File : BE089461.D
 Acq On : 2 Feb 2015 18:31
 Operator : TP/IZ
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.2

Quant Time: Feb 03 03:39:12 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 03 03:06:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.58	152	40363	5.00	ng	0.00
7) Naphthalene-d8	12.49	136	148573	5.00	ng	0.00
13) Acenaphthene-d10	16.68	164	54472	5.00	ng	0.00
17) Phenanthrene-d10	19.99	188	95160	5.00	ng	0.00
20) Chrysene-d12	23.84	240	125834	5.00	ng	0.00
27) Perylene-d12	25.56	264	144870	5.00	ng	0.00

System Monitoring Compounds						
3) 2-Fluorophenol	6.75	112	2778	0.20	ng	0.00
4) Phenol-d6	8.87	99	3202	0.19	ng	0.00
8) Nitrobenzene-d5	10.84	82	2161	0.18	ng	0.00
14) 2,4,6-Tribromophenol	18.59	330	252	0.14	ng	0.00
15) 2-Fluorobiphenyl	15.13	172	3145	0.21	ng	0.00
22) Terphenyl-d14	22.48	244	2872	0.18	ng	0.00

Target Compounds				Qvalue		
2) n-Nitrosodimethylamine	3.68	42	1204m	0.17	ng	
5) Phenol	8.90	94	3454	0.19	ng	96
6) bis(2-Chloroethyl)ether	9.06	93	2545	0.19	ng	# 96
9) Nitrobenzene	10.89	77	2425	0.18	ng	99
10) 2,4-Dimethylphenol	11.83	122	1736	0.18	ng	97
11) 2,4-Dichlorophenol	12.21	162	1344	0.19	ng	82
12) Hexachlorobutadiene	12.87	225	1005	0.21	ng	99
18) Hexachlorobenzene	19.23	284	1057	0.21	ng	92
19) Pentachlorophenol	19.67	266	83	0.09	ng	# 71
21) Benzidine	22.15	184	936	0.11	ng	# 86
23) Benzo(a)anthracene	23.83	228	18740m	0.19	ng	
24) 3,3'-Dichlorobenzidine	23.82	252	1265	0.17	ng	# 92
25) Chrysene	23.87	228	26752	0.21	ng	97
26) Indeno(1,2,3-cd)pyrene	27.07	276	4729m	0.23	ng	
28) Benzo(b)fluoranthene	25.10	252	2214	0.17	ng	# 82
29) Benzo(k)fluoranthene	25.13	252	7875m	0.20	ng	
30) Benzo(a)pyrene	25.48	252	2697	0.15	ng	# 81
31) Dibenzo(a,h)anthracene	27.10	278	2655	0.15	ng	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
Data File : BE089461.D
Acq On : 2 Feb 2015 18:31
Operator : TP/IZ
Sample : SSTDICC0.2
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.2

Quant Time: Feb 03 03:39:12 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 03 03:06:34 2015
Response via : Initial Calibration

