

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
 Data File : BE089463.D
 Acq On : 3 Feb 2015 1:41
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE020215

Quant Time: Feb 03 11:50:42 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 03 11:20:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	n-Nitrosodimethylamine	0.792	0.833	-5.2	77	0.00
3 S	2-Fluorophenol	1.674	1.647	1.6	80	0.00
4 S	Phenol-d6	1.951	1.868	4.3	75	0.00
5 C	Phenol	2.102	1.870	11.0	69	0.02
6	bis(2-Chloroethyl)ether	1.529	1.453	5.0	75	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
8 S	Nitrobenzene-d5	0.363	0.352	3.0	71	0.00
9	Nitrobenzene	0.413	0.411	0.5	71	0.00
10	2,4-Dimethylphenol	0.295	0.296	-0.3	72	0.00
11 C	2,4-Dichlorophenol	0.223	0.224	-0.4	72	0.00
12	Hexachlorobutadiene	0.157	0.152	3.2	72	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.126	0.155	-23.0	85	0.00
15 S	2-Fluorobiphenyl	1.398	1.282	8.3	69	0.00
16 P	2,4-Dinitrophenol	0.033	0.028#	15.2	64	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
18	Hexachlorobenzene	0.262	0.239	8.8	81	0.00
19 C	Pentachlorophenol	0.030	0.024	20.0	82	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
21	Benzidine	0.375	0.457	-21.9	108	0.00
22 S	Terphenyl-d14	0.565	0.605	-7.1	90	0.00
23	Benzo(a)anthracene	3.983	4.116	-3.3	92	0.00
24	3,3'-Dichlorobenzidine	0.354	0.360	-1.7	94	0.00
25	Chrysene	4.866	4.707	3.3	91	0.00
26	Indeno(1,2,3-cd)pyrene	1.440	1.286	10.7	92	0.00
27 I	Perylene-d12	1.000	1.000	0.0	95	0.00
28	Benzo(b)fluoranthene	0.664	0.738	-11.1	98	0.00
29	Benzo(k)fluoranthene	1.246	1.332	-6.9	95	0.00
30 C	Benzo(a)pyrene	0.754	0.835	-10.7	99	0.00
31	Dibenzo(a,h)anthracene	0.915	0.980	-7.1	96	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 0